


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 1366 Accredited to ISO/IEC 17025:2017	ITW Ltd trading as Avery Weigh-Tronix (Calibration Select Laboratory)	
	Issue No: 033	Issue date: 30 October 2023
	Calibration Select Foundry Lane Smethwick West Midlands B66 2LP	Contact: Mr Tom Bates Tel: +44 (0) 121 568 1407 Fax: +44 (0) 121 697 1407 E-Mail: tbates@awtx-itw.com Website: www.calibrationselect.co.uk
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
LOAD CELLS UP TO 50 TONNE CAPACITY	Determination of load cell error, repeatability error and temperature effect on minimum dead load output Determination of creep error Determination of minimum dead load output return Determination of barometric pressure effects Determination of humidity effects Additional tests for load cells equipped with electronics	OIML R60 Edition 2021 (E) OIML R60 Edition 2017 (E) [Withdrawn]Accuracy Class C and D up to a number of verification intervals consistent with the measurement capabilities.
END		