Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

1709

Accredited to ISO/IEC 17025:2017

Tetra Tech Consulting Limited

Issue No: 083 **Issue date:** 27 October 2026

20 Farringdon Street **London** **EC4A 4AB**	**Contact: Chris Parr** **Tel: +44 (0)20 3691 0500** **E-Mail: chris.parr@tetratech.com** **Website: www.tetratech.com**

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 20 Farringdon Street London EC4A 4AB United Kingdom	**Local contact** Chris Parr Tel: +44 (0)20 3691 0500 Email: chris.parr@tetratech.com	Health and Hygiene Head Office Asbestos – All Support Functions	D
	Chris Parr Tel: +44 (0)20 3691 0500 Email: chris.parr@tetratech.com	Legionella Sampling Support Functions: Quality Management, including contract review, document control, auditing and quality control	D

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene Environmental and Water Sampling and Testing	London - K

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	Health and Hygiene	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
ASBESTOS FIBRES IN AIR	Sampling of air for fibre counting	Documented In-House Method RPSCA1 based on HSG 248	K
	Fibre counting	Documented In-House Method RPSCA1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	D, K
	4 Stage Clearance Process	Documented In-House Method RPSCA1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	K
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method RPSCA2 based on HSG 248	K

1709
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Tetra Tech Consulting Limited

Issue No: 083 **Issue date:** 27 October 2026

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS (non regulatory) Drinking Water (mains), Process Water (including Cooling Towers), Recreational Water including Pool Water From sampling location including:- Cooling Towers, Domestic Hot & Cold water systems. Calorifiers, Water dispensers, Cold water tanks, Humidifiers / water, Spa Pools, TMV/mixer taps, Instantaneous water heater, Showers/spray taps and Circulating condenser water (closed open circuit towers)	Water sampling for the purpose of facilities management and as a stand alone practice, for subsequent microbiological analysis: • Aerobic Colony Count • *Legionella* spp. • Coliforms • *Escherichia coli* • *Pseudomonas aeruginosa*	Documented in house method RPSCL/1 based on BS 7592:2022, BS 8554:2015 and The Microbiology of Drinking Water part 2 2010	K
END			