



1810
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Public Health Wales, Microbiology Carmarthenshire
Issue No: 042 Issue date: 05 November 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATERS Drinking (chlorinated, mains and private supplies), Recreational, swimming pools and spa, sea and bathing water, streams and rivers</p>	<p><u>Microbiological Tests</u></p> <p>Enumeration of:</p> <p>Coliforms, confirmed</p> <p><i>Escherichia coli</i>, confirmed</p> <p>Enterococci, confirmed</p> <p>Aerobic micro-organisms at 22 °C and at 37 °C</p>	<p>Documented In-House Methods based on The Microbiology of Drinking Water - Environment Agency</p> <p>W010 based on MDW Part 4A, 2016</p> <p>W010 based on MDW Part 4A, 2016</p> <p>W030 based on MDW Part 5, 2012</p> <p>W020 based on MDW Part 7, 2020</p>
Swimming pool, spa, hydrotherapy pool	<i>Pseudomonas aeruginosa</i>	W040 based on MDW Part 8B, 2015
Endoscope Washer Disinfectors rinse waters	Aerobic Colony Count [30°C] for 5 days	W090 based on BS EN ISO 15883-1:2025
Dialysis water	Aerobic colony count [20°C] for 7 days	W100 based on BS EN ISO 23500-3:2024 and BS EN ISO 23500-5:2024
Drinking, pool/spa waters, hospital and ship waters	<i>Legionella pneumophila</i> serogroups 1 – 14, and other <i>Legionella</i> spp	W080 based on BS EN ISO 11731:2017 Using Matrix A, Procedures 8,9,10, Media A & C
END		