


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|--|--|
|  <p>UKAS TESTING</p> <p>1837</p> <p>Accredited to ISO/IEC 17025:2017</p> | <p>Sheffield Teaching Hospitals NHS Foundation Trust</p> <p>Issue No: 032 Issue date: 18 March 2026</p> | |
| | <p>Northern General Hospital Herries Road Sheffield S5 7BQ</p> | <p>Contact: Richard Flemming Tel: +44 (0) 114 27 15319 Fax: +44 (0)114 242 1385 E-Mail: richard.fleming1@nhs.net</p> |
| <p>Testing performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|---|--|--|
| <p>POOL WATERS POTABLE WATERS DRINKING WATERS</p> <p>(including Hot and Cold Water Systems)</p> | <p><u>Microbiological Tests</u></p> <p><u>Detection & Enumeration:</u></p> <p><i>Legionella</i> spp, confirmed as:</p> <ul style="list-style-type: none"> • <i>Legionella pneumophila</i>, serogroup 1 • <i>Legionella pneumophila</i>, serogroups 2-14 • <i>Legionella</i> spp. <p><u>Enumeration:</u></p> <p>Coliforms, confirmed</p> <p><i>Escherichia coli</i>, confirmed</p> | <p>Documented In-House Method</p> <p>In-House Method BAWA005/006 confirmed using MALDI-TOF-MS in accordance with documented in house method BAAS054 and BAWA005/006 serology procedure using Oxoid latex agglutination identification kit</p> <p>Documented In-House Methods based on The Microbiology of Drinking Water (MDW)</p> <p>BAWA002 based on MDW Part 4a:2016, confirmed using MALDI-TOF-MS in accordance with documented in house method BAAS054</p> <p>BAWA002 based on MDW Part 4a:2016, confirmed using MALDI-TOF-MS in accordance with documented in house method BAAS054</p> |
| <p>Potable Waters, Hospital tap Waters, and Endoscopy rinse waters</p> | <p><i>Pseudomonas aeruginosa</i>, confirmed</p> <p>Total aerobic colony count at 22°C and at 37°C</p> | <p>BAWA004 based on MDW Part 8:2015 and HTM 01-06 Part E (2016) and HTM 04-01 Part B confirmed using MALDI-TOF-MS in accordance with documented in house method BAAS054</p> <p>BAWA003 based on MDW Part 7A:2020</p> |



1837
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Sheffield Teaching Hospitals NHS Foundation Trust
Issue No: 032 **Issue date:** 18 March 2026

Testing performed at main address only

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|---|--|---|
| POOL WATERS POTABLE WATERS DRINKING WATERS (cont'd) Endoscopy rinse waters | <u>Microbiological Tests</u> (cont'd) <u>Enumeration:</u> (cont'd) Total Viable Count at 30°C for 5 Days | Documented In-House Method BAWA008 based on HTM-01-06 Part E (2016). |
| END | | |