


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	AWE PLC trading as AWE Nuclear Security Technologies	
	Issue No: 015	Issue date: 27 May 2025
	Building A14.1 Aldermaston Reading Berks RG7 4PR	Contact: No Commercial Enquiries Tel: Fax: E-Mail: Website:
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Building A14.1 Aldermaston Reading Berks RG7 4PR	Local contact No Commercial Enquiries Health and Hygiene Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



1854
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

AWE PLC trading as AWE Nuclear Security Technologies

Issue No: 015 Issue date: 27 May 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method LQP007 based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected on containing asbestos	Fibre counting	Documented In-House Method LQP007, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method LQP003 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method LQP004 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			