Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

1995

Accredited to ISO/IEC 17025:2017

SUEZ Recycling and Recovery UK Ltd

Issue No: 028 **Issue date:** 06 August 2026

Tees Valley Laboratory
Catherine Road
New Herrington Industrial Estate
Houghton-le-Spring
DH4 7BG

Contact: Mr Alan Shepherd
Tel: +44 (0)1642 371 471
E-Mail: alan.shepherd@suez.com
Website: www.suez.co.uk

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
Incinerator Bottom Ash (IBA)	<u>Chemical and Physical Tests</u>	Documented in-house methods
	pH Alkali Reserve	IBA01 by probe and titrimetry
	The preparation (drying and grinding) of IBA for subsequent analysis by IBA01 and IBA02	Methods IBA03 – IBA07
Incinerator Bottom Ash (IBA)	Metals: Arsenic Cadmium Cobalt Chromium Copper Nickel Manganese Lead Antimony Vanadium Zinc	IBA19 using Aqua Regia Digestion and ICP OES
Incinerator Bottom Ash (IBA)	Metals: Mercury Thallium	IBA21 using Aqua Regia Digestion and ICPMS
Incinerator Bottom Ash (IBA)	Total Organic Carbon	IBA24 by TOC Infra-red Analyser
Air Pollution Control Residue (APCR)	Metals: Mercury Thallium	LIM03 using Aqua Regia Digestion and ICPMS

1995

Accredited to ISO/IEC 17025:2005

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SUEZ Recycling and Recovery UK Ltd

Issue No: 028 **Issue date:** 06 August 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
	<u>Chemical and Physical Tests</u>	Documented in-house methods
Air Pollution Control Residue (APCR)	Metals: Arsenic Cadmium Cobalt Chromium Copper Nickel Manganese Lead Antimony Vanadium Zinc	LIM02 using Aqua Regia Digestion and ICPOES
Waste materials including Trommel Fines	Determination of Loss on ignition at 440°C	LOI01 by Gravimetic analysis in accordance with HMRC Excise Notice LFT1 21st September 2023
END		