


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2054</b> <b>Accredited to ISO/IEC 17025:2017</b>	<b>Airborne Environmental Consultants Limited</b>	
	<b>Issue No:</b> 023 <b>Issue date:</b> 20 November 2024	
	<b>23 &amp; 27 Wheelforge Way</b> <b>Ashburton Point</b> <b>Trafford Park</b> <b>Manchester</b> <b>M17 1EH</b>	<b>Contact: Mr Daniel Shuttleworth</b> <b>Tel: +44 (0)161 872 7111</b> <b>E-Mail: sales@aec.uk</b> <b>Website: www.aec.uk</b>
<b>Testing performed by the Organisation at the locations specified below</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details		Activity	Location code
<b>Address</b> 23 & 27 Wheelforge Way Ashburton Point Trafford Park Manchester M17 1EH	<b>Local contact</b> Mr Daniel Shuttleworth  <b>Tel: 0161 872 7111</b>	Health and Hygiene Asbestos – All Support Functions	A
<b>Address</b> Fernwood House The Hilway Mountnessing Brentwood CM15 0UE	<b>Local contact</b> Mr Daniel Shuttleworth  <b>Tel: 0203 384 6175</b>	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none"><li>Contract Review</li><li>Reporting</li></ul>	C

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	D



2054  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Airborne Environmental Consultants Limited**  
**Issue No: 023   Issue date: 20 November 2024**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method AEC 1 based on HSG 248	B
	Fibre counting	Documented In-House Method AEC 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, D
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method AEC 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method AEC 6 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method AEC 2 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, C
END			