Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

2054

Accredited to ISO/IEC 17025:2017

Airborne Environmental Consultants Limited

Issue No: 024 **Issue date:** 11 August 2026

23 Wheelforge Way
Ashburton Point
Trafford Park
Manchester
M17 1EH

Contact: Mr Daniel Shuttleworth
Tel: +44 (0)161 872 7111
E-Mail: sales@aec.uk
Website: www.aec.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address 23 Wheelforge Way Ashburton Point Trafford Park Manchester M17 1EH	**Local contact** Mr Daniel Shuttleworth Tel: 0161 872 7111	Health and Hygiene Asbestos – All Support Functions	A
Address Fernwood House The Hilway Mountnessing Brentwood CM15 0UE	**Local contact** Mr Daniel Shuttleworth Tel: 0203 384 6175	Health and Hygiene Asbestos – Support Functions: • *Contract Review* • *Reporting*	C

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	D

2054
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Airborne Environmental Consultants Limited

Issue No: 024 **Issue date:** 11 August 2026

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	Health and Hygiene	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
ASBESTOS FIBRES IN AIR	Sampling of air for fibre counting	Documented In-House Method AEC 1 based on HSG 248	B
	Fibre counting	Documented In-House Method AEC 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, D
	4 Stage Clearance Process	Documented In-House Method AEC 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method AEC 6 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method AEC 2 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, C
END			