


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>2136</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<h3>APEC Environmental Limited</h3> <p>Issue No: 024 Issue date: 17 June 2025</p>	
	<p>7 Apollo Court Koppers Way Monkton Business Park South Hebburn Tyne and Wear NE31 2ES</p>	<p>Contact: Mr C Watson Tel: +44 (0)1142 459371 Fax: +44 (0)1709 363642 Email: chris@apecuk.com Website: www.apecuk.com</p>
<p>Testing performed by the Organisation at the locations specified below</p>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<p><b>Address</b> 7 Apollo Court Koppers Way Monkton Business Park South Hebburn Tyne and Wear NE31 2ES</p> <p><b>Local contact</b> Mr P Todd Tel: +44 (0)191 438 5884 Fax: +44 (0)191 438 6375 Email: phil@apecuk.com</p>	<p>Health and Hygiene Head Office</p> <p>Asbestos – All Support Functions</p>	A
<p><b>Address</b> 9 Telford Court Dunkirk Lea Chester Gate Chester CH1 6LT</p> <p><b>Local contact</b> Mr D Mills Tel: +44 (0)1244 851753 Fax: +44 (0)1244 851752 Email: danny.mills@apecuk.com</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	B
<p><b>Address</b> Office 5, Haywood House Hydra Business Park Nether Lane Ecclesfield Sheffield S35 9ZX</p> <p><b>Local contact</b> Mr C Watson Tel: +44 (0)1142 459371 Fax: +44 (0)1709 363642 Email: chris@apecuk.com</p>	<p>Health and Hygiene Asbestos – Support Functions:</p> <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	C



2136  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**APEC Environmental Limited**  
**Issue No: 024 Issue date: 17 June 2025**

Testing performed by the Organisation at the locations specified

<b>Address</b> 2 <sup>nd</sup> Floor Spinnaker House 48 High Street Bagshot Surrey GU19 5AW	<b>Local contact</b> Mr M Curtis  Tel: +44 (0) 1276 472596 Fax: +44 (0) 1276 479182 Email: michael@apecuk.com	Health and Hygiene Asbestos – Support Functions: <ul style="list-style-type: none"> <li>• <i>Quality Audit Administration</i></li> <li>• <i>Contract Review</i></li> <li>• <i>Scheduling</i></li> <li>• <i>Personnel</i></li> <li>• <i>Equipment</i></li> <li>• <i>Measurement Traceability</i></li> <li>• <i>Reporting</i></li> </ul>	D
--	--	---	---

**Site activities performed away from the locations listed above:**

Location details	Activity	Location code
Client Premises                      Hebburn	Health and Hygiene	E
Client Premises                      Chester	Health and Hygiene	F
Client Premises                      Ecclesfield	Health and Hygiene	G
Client Premises                      Bagshot	Health and Hygiene	H



2136  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**APEC Environmental Limited**  
**Issue No: 024 Issue date: 17 June 2025**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method OH 1 based on HSG 248	A, B, C, D, E, F, G, H
	Fibre counting	Documented In-House Method OH 2, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C, D, E, F, G, H
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method OH 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	E, F, G, H
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method OH 3 based on HSG 248 Documented In-House Method OH 4 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	E, F, G, H
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite		A, B, C, D
END			