


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 21701 Accredited to ISO/IEC 17025:2017	Smithers MDT Limited	
	Issue No: 004 Issue date: 18 November 2022	
	Shawbury Shrewsbury Shropshire SY4 4NR	Contact: Richard Barlow Tel: +44 (0)1939 250383 E-Mail: rbarlow@smithers.com Website: www.smithers.com
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
MEDICAL DEVICE TESTING		
Vials	Container closure seal integrity	BS EN ISO 8871-5:2016 (Annex D)
Pre-Filled Syringes	Breakloose and glide force	BS ISO 11040-4:2015 (Annex E) BS ISO 11040-8:2016
	Dye solution tightness test	BS ISO 11040-4:2015 + A1:2020 (Annex H)
Needle based injection systems	Dose accuracy	BS EN ISO 11608-1:2015 BS EN ISO 11608-1:2022
Syringes	Plunger operation force	BS EN ISO 7886-1:2018 (Annex E) BS EN ISO 7886-1:1997 (Annex G)
END		