


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION</p> <p>21817</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Thermosense Limited</p> <p>Issue No: 007 Issue date: 20 June 2026</p>	
	<p>Eton House Eton Way North Radcliffe Lancashire M26 2ZT</p>	<p>Contact: Mr Ryan Davis Tel: +44 (0)162 8531 166 E-Mail: ryan.davis@thermosense.co.uk Website: www.thermosense.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
TEMPERATURE			
Base Metal thermocouples (standalone and/or with indicators)	-80 °C to -45 °C -45 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.90 °C 0.50 °C 0.50 °C 0.47 °C 2.8 °C	Calibration in metal block baths
Noble Metal thermocouples Type R and S	0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.54 °C 0.59 °C 2.8 °C	Calibration in metal block baths
Probes and indicators	-80 °C to -45 °C -45 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.80 °C 0.30 °C 0.30 °C 0.31 °C 2.8 °C	Calibration in metal block baths
4-wire Platinum Resistance Thermometers	-80 °C to -45 °C -45 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C	0.80 °C 0.20 °C 0.20 °C 0.46 °C	Calibration in metal block baths. 3-wire probes will be subject to additional uncertainty contributions.
Probes/thermocouples with 4-20 mA Transmitters	-80 °C to -45 °C -45 °C to 0 °C 0 °C to 100 °C 100 °C to 650 °C 650 °C to 1200 °C	0.80 °C 0.30 °C 0.30 °C 0.33 °C 2.8 °C	Calibration in metal block baths
Temperature indicators with sensors in air	5 °C to 50 °C	0.40 °C	By comparison with reference PT100 thermometers. Calibration of devices with analogue outputs may be undertaken.
Temperature by Electrical Simulation			
Pt1000 simulation	-200 °C to +850 °C	0.034 °C	
Resistance thermometers (Pt100)	-200 °C to +850 °C	0.019 °C	



21817

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thermosense Limited
Issue No: 007 Issue date: 20 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
TEMPERATURE (continued)			
Thermocouple Simulation			Including cold junction compensation
Type B	300 to 600 °C 600 to 1820 °C	0.50 °C 0.26 °C	
Type E	-210 to -100 -100 to 1000	0.50 °C 0.26 °C	
Type J	-210 °C to -100 °C -100 °C to 1200 °C	0.37 °C 0.18 °C	
Type K	-210 °C to -100 °C -100 °C to 1300 °C	0.47 °C 0.22 °C	
Type N	-210 °C to -100 °C -100 °C to 1300 °C	0.64 °C 0.27 °C	
Type R	0 °C to 600 °C 600 °C to 1700 °C	0.40 °C 0.20 °C	
Type S	0 °C to 600 °C 600 °C to 1700 °C	0.39 °C 0.22 °C	
Type T	-210 °C to -100 °C -100 °C to 400 °C	0.42 °C 0.22 °C	
HUMIDITY			
Relative humidity instruments	5 °C to 10 °C 10 %rh to 50 %rh 50 %rh to 90 %rh	0.40 °C 1.7 %rh 2.3 %rh	By comparison with reference thermo-hygrometers. Calibration of devices with analogue outputs may be undertaken (4-20mA transmitters and voltage transducers)
	10 °C to 40 °C 5 %rh to 50 %rh 50 %rh to 90 %rh	0.40 °C 1.5 %rh 1.9 %rh	
	40 °C to 50 °C 10 %rh to 50 %rh 50 %rh to 90 %rh	0.40 °C 1.4 %rh 1.7 %rh	
END			



21817

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thermosense Limited
Issue No: 007 Issue date: 20 June 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$