

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

| | | |
|--|---|---|
|  <p>UKAS TESTING</p> <p>21871</p> <p>Accredited to ISO/IEC 17025:2017</p> | <h3>Cambridge Acoustic Laboratory Limited</h3> <p>Issue No: 005 Issue date: 18 December 2025</p> | |
| | <p>Brewery Road Pampisford Cambridge CB22 3HG</p> | <p>Contact: Sophia McNally Tel: +44 (0)1223 837007 E-Mail: testing@camacousticlab.co.uk Website: www.camacousticlab.co.uk</p> |
| <p>Testing performed at the above address only</p> | | |

DETAIL OF ACCREDITATION

| Materials/Products tested | Type of test/Properties measured/Range of measurement | Standard specifications/ Equipment/Techniques used |
|--|--|--|
| <p><u>Building Elements</u> Building elements including doors, windows, glazing, ventilators, grilles, cable raceways, shutters, etc</p> <p>Doors BS EN ISO 10140-1: 2021 (Annex B) Windows BS EN ISO 10140-1: 2021 (Annex C) Glazing BS EN ISO 10140-1: 2021 (Annex D) Small technical elements such as ventilators, grills and cable raceways BS EN ISO 10140-1: 2021 (Annex E) Shutters BS EN ISO 10140-1: 2021 (Annex I) Joints filled with fillers or seals BS EN ISO 10140-1:2021 (Annex J) Building elements and facades not specifically described in the Annexes of BS EN ISO 10140-1: 2021</p> | <p>Laboratory measurement of airborne sound insulation of building elements</p> <p>The property measured is the Sound Reduction Index over the minimum frequency range of 100 Hz - 3150 Hz</p> | <p>BS EN ISO 10140-2:2021 BS EN ISO 717-1:2020</p> |
| <p>END</p> | | |