

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS TESTING 2210 Accredited to ISO/IEC 17025:2017	SembCorp Utilities (UK) Limited	
	Issue No: 047 Issue date: 04 October 2024	
	Analytical Services Building Wilton International Middlesbrough TS90 8WS	Contact: Euan Cramond Tel: +44 (0)1642 212373 Fax: +44 (0)1642 212718 E-Mail: euan.cramond@sembcorp.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Analytical Services Building Wilton International Middlesbrough TS90 8WS Local contact Mr Euan Cramond Tel: +44 (0)1642 212373 Fax: +44 (0)1642 212718 Email: euan.cramond@sembcorp.com	Support functions: Quality System Quality Audit Administration Sampling and Testing: Health and Hygiene	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



2210
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

SembCorp Utilities (UK) Limited
Issue No: 047 Issue date: 04 October 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method DI/AN/ME/025 based on HSG 248	B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	Fibre counting	Documented In-House Method DI/AN/ME/001 based on HSG 248	A, B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method DI/AN/ME/026 based on HSG 248	A
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method DI/AN/ME/003 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	
END			