

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2230 Accredited to ISO/IEC 17025:2017	Asbestos Analytical Services Limited Trading As Asbestos Analytical Services	
	Issue No: 018 Issue date: 03 May 2024	
Unit 1 Shawfarm Road Prestwick KA9 2TR	Contact: Mr D Espie Tel: +44 (0)1292 475176 Fax: +44 (0)1292 474845 E-Mail: info@aas-ayr.co.uk Website: aasayr.co.uk	
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code	
Address Unit 1 Shawfarm Road Prestwick KA9 2TR	Local contact Mr D Espie	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B



2230
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Asbestos Analytical Services Limited Trading As Asbestos Analytical Services
Issue No: 018 Issue date: 03 May 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method LSD.ASB.02 based on HSG 248	A, B
	Fibre counting	Documented In-House Method LSD.ASB.02, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method LSD.ASB.02, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for subsequent identification of asbestos	Documented In-House Method LSD.ASB.04 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method LSD.ASB.01 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A

END