


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 Accredited to ISO/IEC 17025:2017	<b>Medical, Health and Education Limited</b>	
	<b>Issue No:</b> 001	<b>Issue date:</b> 31 July 2024
	<b>46 – 48 Grove Street</b> <b>Swadlincote</b> <b>Derbyshire</b> <b>DE11 9DD</b>	<b>Contact: Christopher Evans</b> <b>Tel: +44 (0)128 3375 093</b> <b>E-Mail: info@mhe.ltd</b> <b>Website: www.mhe.ltd</b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUIDS	<u>Workplace Drugs Analysis/Medical and Legal Analysis</u>	Documented In-House Methods:
Urine (unpreserved)	Adulteration tests: pH Creatinine Specific Gravity	SOP005 using Siemens Clinitek Status+
Urine (unpreserved)	Confirmation and Quantitative analysis of the following drugs (cut-off limit/s); [Concentration Range]:  <b>Amphetamines group:</b> Amphetamine (200 ng/ml) [40-1600 ng/ml] MDA (200 ng/ml); [40-1600 ng/ml] MDMA (200 ng/ml); [40-1600 ng/ml] Methamphetamine (200 ng/ml); [40-1600 ng/ml]  <b>Benzodiazepine group:</b> Alprazolam (100 ng/ml); [20-800 ng/ml] Diazepam (100 ng/ml); [20-800 ng/ml] Lorazepam (100 ng/ml); [20-800 ng/ml] Nordiazepam (100 ng/ml); [20-800 ng/ml] Oxazepam (100 ng/ml); [20-800 ng/ml] Temazepam (100 ng/ml); [20-800 ng/ml]	SOP007 using Sample dilution and LC-MS/MS  SOP007 using Sample dilution and LC-MS/MS

