


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 23364 Accredited to ISO/IEC 17025:2017	Werth Metrology Limited Issue No: 003 Issue date: 24 April 2026	
	35 Brunel Parkway Derby DE24 8HR	Contact: Carl Harrison Tel: +44 (0) 1332 289664 E-Mail: service@werth-metrology.com Website: www.werth-metrology.com

Calibration performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 35 Brunel Parkway Derby DE24 8HR	Administrative activities only	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The location must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer	Dimensional	B



23364
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Werth Metrology Limited
Issue No: 003 Issue date: 24 April 2026

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k=2$)	Remarks	Location Code
MEASURING INSTRUMENTS AND MACHINES Cartesian co-ordinate measuring machines (CMM's)	Length measurement: E_L 0 to 635 mm (Longest diagonal) Single stylus probing test: $P_{Form.Sph.1 \times 25:SS:Tact}$ $P_{Size.Sph.1 \times 25:SS:Tact}$	0.25 + (5.0 x length in metres) μm 0.84 μm 0.68 μm	ISO 10360-2:2009 using end standards ISO 10360-5:2020 Using a 10 mm to 51 mm diameter test sphere. Test value uncertainties calculated in line with ISO/TS 17865:2016. Accreditation is limited to CMM's manufactured by Werth.	B
END				



23364
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Werth Metrology Limited
Issue No: 003 Issue date: 24 April 2026

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$