

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p><b>23543</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Kingston University Enterprise Limited</b> trading as <b>Kingston Analytical Services Toxicology (KAST)</b></p> <p><b>Issue No: 007      Issue date: 06 April 2026</b></p>	
	<p><b>Hind Court</b> 106 - 114 London Road Kingston KT2 6TN GBR</p>	<p><b>Contact: Dr B Rooney</b> Tel: +44 (0)208 417 2048 E-Mail: <a href="mailto:KAST@kingston.ac.uk">KAST@kingston.ac.uk</a> Website: <a href="http://www.kastforensics.com/">www.kastforensics.com/</a></p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS</p> <p>Whole Blood (Preserved)</p>	<p>Detection and quantitation of drugs in relation to s5A of the Road Traffic Act 1988 (as amended) and The Drug Driving (Specified Limits) (England and Wales) Regulations 2014 (Cut-off) [Concentration Range]:</p> <ul style="list-style-type: none"> <li>- Benzoylcegonine (50 µg/L) [25-500ng/mL]</li> <li>- Cocaine (10 µg/L) [5-100ng/mL]</li> <li>- Delta-9-tetrahydrocannabinol (2 µg/L) [1-20ng/mL]</li> <li>- Morphine (80 µg/L); [40-140 µg/L]</li> <li>- 6-AM (5 µg/L); [2.5-15 µg/L]</li> <li>- Methadone (500 µg/L); [250-875 µg/L]</li> </ul>	<p>The organisation has demonstrated compliance to the Forensic Science Regulator Code of Practice V2 in relation to the Forensic Activities listed below.</p> <p>In addition, where compliance has been demonstrated for the related FSA specific requirements this is stated below at the relevant schedule entry.</p> <p>The organisation has demonstrated compliance to the Forensic Science Regulator Code of Practice V2 FSA Specific Requirements: Toxicology: analysis for drugs in relation to s5A of the Road Traffic Act 1988</p> <p>Documented in house methods:</p> <p>KSTPr 013 by GCMS-MS</p> <p>KSTPr 014 by GCMS-MS</p> <p>KSTPr 053 by GCMS-MS</p>



23543

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Kingston University Enterprise Limited t/a Kingston Analytical Services Toxicology (KAST)

Issue No: 007 Issue date: 06 April 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUIDS (cont'd)	<ul style="list-style-type: none"><li>-Amphetamine (250 µg/L); [50-500 µg/L]</li><li>- Methamphetamine (10 µg/L); [5-100 µg/L]</li><li>- MDMA (10 µg/L); [5-100 µg/L]</li><li>- Diazepam (550 µg/L); [250-1000 µg/L]</li><li>- Lorazepam (100 µg/L); [50-200 µg/L]</li><li>- Flunitrazepam (300 µg/L); [100-600 µg/L]</li><li>- Temazepam (1000 µg/L); [200-1600 µg/L]</li><li>- Ketamine (20 µg/L); [5-100 µg/L]</li><li>- Clonazepam (50 µg/L); [10-100 µg/L]</li><li>- Oxazepam (300 µg/L); [100-600 µg/L]</li></ul>	<p>KSTPr 052 by LCMSMS</p> <p>The organisation has demonstrated compliance to the Forensic Science Regulator Code of Practice V2 in relation to the Forensic Activities listed below:</p>

