


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS CALIBRATION 23653</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Thorcal Ltd</p> <p>Issue No: 011 Issue date: 30 January 2026</p>	
	<p>B6 Citadel Trading Park Hull HU9 1TQ</p>	<p>Contact: James Thompson Tel: +44 (0) 1482 242010 E-Mail: Info@thorcal.co.uk Website: www.thorcal.co.uk</p>
<p>Calibration performed at the above address only</p>		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
<p>PRESSURE</p> <p><u>Hydraulic pressure (gauge)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>275 kPa to 690 kPa 690 kPa to 5.5 MPa 5.5 MPa to 110 MPa</p>	<p>0.012 % 0.010 % 0.0080 %</p>	<p>Methods consistent with EURAMET CG17</p> <p>Calibration of pressure measuring devices with an electrical output may be undertaken.</p> <p>Calibration on other fluids up to 70 MPa can be undertaken with an additional uncertainty of 1.4 kPa added in quadrature</p>
<p><u>Gas pressure (gauge)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>-100 kPa to -6.9 kPa -6.9 kPa to 0 Pa 0 Pa to 1.5 kPa 1.5 kPa to 2.5 kPa 2.5 kPa to 10 kPa 10 kPa to 12.15 MPa</p>	<p>0.0080 % 5.6 Pa 11 Pa Q [0.0060 %, 0.30 Pa] Q [0.0060 %, 0.15 Pa] 0.0075 %</p>	<p>Absolute pressure calibrations may be undertaken by associated barometric pressure measurement with an additional uncertainty of 13 Pa added in quadrature.</p>
<p><u>Gas pressure (absolute)</u></p> <p>Calibration of pressure indicating instruments and gauges</p>	<p>75 kPa to 115 kPa</p>	<p>10 Pa to 13 Pa</p>	
<p>TEMPERATURE</p> <p>Temperature indicators and or recorders with temperature sensors</p>	<p>-45 °C to 0 °C 0 °C to 75 °C 75 °C to 232 °C 232 °C to 420 °C 420 °C to 660 °C</p>	<p>0.13 °C 0.12 °C 0.24 °C 0.28 °C 0.42 °C</p>	<p>By comparison in a range of metal and liquid media baths</p>
<p>Metal Block Calibrators</p>	<p>-90 °C to 0 °C 0 °C to 125 °C 125 °C to 660 °C</p>	<p>0.040 °C 0.030 °C 0.030 °C</p>	<p>Comparison to a reference PRT. Characterisation consistent with Euramet CG 13</p>



23653
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd
Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
TEMPERATURE (continued)			
Platinum Resistance Thermometers	-39 °C to 0 °C 0.01 °C 0 °C to 232 °C 232 °C to 420 °C 420 °C to 660 °C	0.10 °C 0.050 °C 0.26 °C 0.27 °C 0.43 °C	3 and 4 wire PRTs Output of the PRTs 4-20mA transmitters.
Thermocouple (Base metal)	-39 °C to 0 °C 0.01 °C 0 °C to 232 °C 232 °C to 420 °C 420 °C to 660 °C	0.30 °C 0.29 °C 0.47 °C 0.62 °C 0.92 °C	
ELECTRICAL			All electrical calibrations are performed as a comparison against a reference standard
DC VOLTAGE			
Generation	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1050 V	46 µV/V + 2.1 µV 46 µV/V + 31 µV 46 µV/V + 310 µV 47 µV/V + 3.7 mV 69 µV/V + 35 mV	For the calibration of measuring instruments
Measurement	0 to 100 mV 100 mV to 1 V 1 V to 10 V 10 V to 100 V 100 V to 1000 V	9.2 µV/V + 1.6 µV 1.7 µV/V + 5.8 µV 1.5 µV/V + 53 µV 4.9 µV/V + 580 µV 47 µV/V + 5.8 mV	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
DC CURRENT			
Generation	0 A to 200 µA 200 µA to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A 2 A to 20 A	0.018 % + 30 nA 0.014 % + 87 nA 0.014 % + 640 nA 0.014 % + 11 µA 0.046 % + 200 µA 0.070 % + 10 mA	For the calibration of measuring instruments
Simulated Current	10 A to 100 A 100 A to 1100 A	0.001 % + 1.9 A 0.0085 % + 1.9 A	For the calibration of clamp meters, using multi-turn coil technique.
Measurement	0 µA to 100 µA 100 µA to 1 mA 1 mA to 10 mA 10 mA to 100 mA 100 mA to 1 A 1 A to 3 A	28 µA/A + 3.4 nA 20 µA/A + 21 nA 140 µA/A + 0.21 µA 63 µA/A + 2.4 µA 10 µA/A + 48 µA 0.15 % + 0.43 mA	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.



23653

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd

Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL (continued) AC VOLTAGE			
Generation	<u>1 mV to 200 mV</u> 40 Hz to 1 kHz 1 kHz to 20 kHz	0.058% + 0.12 mV 0.12 % + 0.18 mV	For the calibration of measuring instruments Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
	<u>0.2 V to 2 V</u> 40 Hz to 1 kHz 1 kHz to 20 kHz	0.046 % + 0.33 mV 0.069 % + 0.60 mV	
	<u>2 V to 20 V</u> 40 Hz to 1 kHz 1 kHz to 20 kHz	23 mV 21mV	
	<u>20 V to 200 V</u> 40 Hz to 1 kHz	0.077% + 42 mV	
	<u>200 V to 1000 V</u> 40 Hz to 1 kHz	0.10 % + 14 mV	
Measurement	10 mV to 100 mV 10 Hz to 60 Hz 60 Hz to 1 kHz 1 kHz to 50 kHz	72 μ V/V + 39 μ V 84 μ V/V + 19 μ V 51 μ V/V + 50 μ V/V	
	100 mV to 1 V 10 Hz to 60 Hz 60 Hz to 1 kHz 1 kHz to 50 kHz	81 μ V/V + 0.14 mV 67 μ V/V + 0.07 mV 97 μ V/V + 1.5 mV	
	1 V to 10 V 10 Hz to 60 Hz 60 Hz to 1 kHz 1 kHz to 100 kHz 100 kHz to 1 MHz	70 μ V/V + 1.2 mV 68 μ V/V + 0.62 mV 82 μ V/V + 1.9 mV 0.17 % + 63 mV	
	10 V to 100 V 40 Hz to 1 kHz 1 kHz to 10 kHz 10 kHz to 50 kHz	25 μ V/V + 12 mV 27 μ V/V + 5.6 mV 170 μ V/V + 5.2 mV	
	100 V to 700 V 40 Hz to 10 kHz	0.02 % + 50 mV	



23653

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd

Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	
ELECTRICAL (continued)				
AC CURRENT				
Generation	<u>0 A to 200 μA</u> 40 Hz to 1 kHz	0.081 % + 0.37 μ A	For the calibration of measuring instruments	
	<u>200 μA to 2 mA</u> 40 Hz to 1 kHz	0.11 % + 0.88 μ A		
	<u>2 mA to 20 mA</u> 40 Hz to 1 kHz	0.081 % + 8.7 μ A		
	<u>20 mA to 200 mA</u> 40 Hz to 1 kHz	0.092 % + 87 μ A		
	<u>200 mA to 2 A</u> 40 Hz to 500 Hz	0.12 % + 1.6 mA		
	<u>2 A to 20 A</u> 40 Hz to 500 Hz	0.23 % + 7.1 mA		
Measurement	0 μ A to 100 μ A 10 Hz to 1 kHz	83 μ A/A + 0.12 μ A		
	100 μ A to 1 mA 10 Hz to 1 kHz	19 μ A/A + 0.12 μ A		
	1 mA to 10 mA 20 Hz to 1 kHz 1 kHz to 5 kHz	49 μ A/A + 2.8 μ A 75 μ A/A + 5.4 μ A		
	10 mA to 100 mA 10 Hz to 1 kHz	46 μ A + 28 μ A		
	100 mA to 1 A 10 Hz to 1 kHz	49 μ A/A + 0.44 mA		
	1 A to 3 A	13 mA		
Simulated Current	10 A to 100 A 40 Hz to 1 kHz	0.0014 % + 1.9 A		For the calibration of clamp meters, using multi-turn coil technique.
	100 A to 1100 A 40 Hz to 100 Hz	0.016 % + 1.9 A		



23653

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd

Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
ELECTRICAL (continued)			
DC RESISTANCE			
FIXED RESISTANCE 2 WIRE or 4 WIRE			
Generation	10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω 1 M Ω 10 M Ω 100 M Ω 1 G Ω	13 m Ω 17 m Ω 79 m Ω 0.79 Ω 7.9 Ω 0.18 k Ω 12 k Ω 1.2 M Ω 110 M Ω	For the calibration of measuring instruments
Measurement	0 Ω to 10 Ω 0 Ω to 100 Ω 100 Ω to 1 k Ω 1 k Ω to 10 k Ω 10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω	29 $\mu\Omega/\Omega$ + 3.5 m Ω 8.6 $\mu\Omega/\Omega$ + 3.6 m Ω 4.3 $\mu\Omega/\Omega$ + 8.9 m Ω 1.9 $\mu\Omega/\Omega$ + 77 m Ω 8.8 $\mu\Omega/\Omega$ + 1.2 Ω 37 $\mu\Omega/\Omega$ + 24 Ω 8.9 $\mu\Omega/\Omega$ + 0.45 k Ω 0.02 % + 36 k Ω 0.35 % + 1.1 M Ω	Generation of these quantities with the same or similar CMCs may be undertaken over the same ranges by the use of a transfer method.
SIMULATED RESISTANCE	0 Ω to 400 Ω 400 Ω to 4000 Ω 4000 Ω to 40 k Ω	0.058 % + 0.58 Ω 0.027 % + 1.2 Ω 0.023 % + 5.9 Ω	For the calibration of measuring instruments



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd
Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks
Temperature indicators, calibration by electrical simulation			
Type J Thermocouple	-210 °C to 50 °C 50 °C to 1200 °C	0.37 °C 0.29 °C	Excluding internal reference junction (CJC)
Type K Thermocouple	-200 °C to 0 °C 0 °C to 1372 °C	0.44 °C 0.36 °C	
Type T Thermocouple	-200 °C to 100 °C 100 °C to 400 °C	0.42 °C 0.27 °C	
Type N Thermocouple	-200 °C to 0 °C 0 °C to 600 °C 600 °C to 1300 °C	0.26 °C 0.29 °C 0.34 °C	
Type E Thermocouple	-200 °C to 0 °C 0 °C to 1000 °C	0.31 °C 0.27 °C	
Type R Thermocouple	-50 °C to 50 °C 50 °C to 250 °C 250 °C to 1768 °C	2.0 °C 0.87 °C 0.64 °C	
Type S Thermocouple	-50 °C to 500 °C 500 °C to 1768 °C	1.5 °C 0.67 °C	
Type B Thermocouple	300 °C to 800 °C 800 °C to 1820 °C	2.0 °C 0.89 °C	
PT100	-39 °C to 0 °C 0 °C to 420 °C 420 °C to 660 °C	0.15 °C 0.30 °C 0.50 °C	
18th EDITION ELECTRICAL TEST EQUIPMENT			
CONTINUITY RESISTANCE	0 Ω to 1 Ω 1 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 kΩ	0.29 % + 31 mΩ 0.29 % + 31 mΩ 0.29 % + 31 mΩ 0.29 % + 31 mΩ	For the calibration of Insulation Resistance Meters using a dedicated calibration system
CONTINUITY CURRENT	10 mA to 320 mA	1.5 % + 0.7 mA	
INSULATION RESISTANCE	10 kΩ to 100 kΩ 100 kΩ to 1 MΩ 1 MΩ to 10 MΩ 10 MΩ to 100 MΩ 100 MΩ to 1 GΩ 1 GΩ to 10 GΩ	0.12 % + 0.20 Ω 0.12 % 1.2 % 1.2 % 1.4 % 6.5 %	



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd
Issue No: 011 Issue date: 30 January 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
18th EDITION ELECTRICAL TEST EQUIPMENT (continued)			
INSULATION TEST VOLTAGE	50 V to 1000 V	1.2 % + 0.93 V	
AC VOLTAGE OUTPUT	100 V to 200 V 200 V to 400 V At 50 Hz	0.23 % + 0.16 V 0.23 % + 0.15 V	
RCD CURRENT	At 50 Hz <u>20 ms to 190 ms</u> 2 mA to 1 A 1 A to 3 A	5.8 % + 0.23 mA 5.8 % + 0.23 mA	For the calibration of RCD Testers using a dedicated calibration system
	<u>190 ms to 5 s</u> 2 mA to 1 A 1 A to 3 A	1.4 % + 71 uA 1.4 % + 71 uA	
TRIP TIME	20 ms to 400 ms 400 ms to 5 s	0.95 ms 8.1 ms	
LOOP RESISTANCE	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1 k Ω At 50 Hz	0.58 % + 41 m Ω 0.58 % + 45 m Ω 0.58 % + 54 m Ω	For the calibration of Earth Loop Testers using a dedicated calibration system
PAT TESTERS			
PAT EARTH BOND RESISTANCE at 50 Hz	0 Ω to 10 Ω 10 Ω to 100 Ω 100 Ω to 1000 Ω	0.58 % + 4.7 m Ω 0.58 % + 19 m Ω 0.58 % + 35 m Ω	For the calibration of Portable Appliance Testers using a dedicated calibration system
PAT EARTH BOND CURRENT at 50 Hz	0 mA to 500 mA 500 mA to 10 A 10 A to 30 A	1.7 % + 14 mA 1.7 % + 71 mA 1.7 % + 71 mA	
PAT INSULATION RESISTANCE	10 k Ω to 100 k Ω 100 k Ω to 1 M Ω 1 M Ω to 10 M Ω 10 M Ω to 100 M Ω 100 M Ω to 1 G Ω 1 G Ω to 10 G Ω	0.12 % + 0.20 Ω 0.12 1.2 % 1.2 % 1.4 % 6.5 %	
PAT LEAKAGE CURRENT at 50 Hz	2.7 mA at 240 V 4.7 mA at 240 V 7.7 mA at 240 V	1.7 % + 9.6 μ A 1.7 % + 9.6 μ A 1.7 % + 9.6 μ A	
END			



23653

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Thorcal Ltd

Issue No: 011 **Issue date:** 30 January 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$