

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>2450</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p>Mercian Science Ltd</p> <p>Issue No: 033 Issue date: 05 December 2025</p>	
	<p>Aqua House Europa Way Britannia Enterprise Park Lichfield Staffordshire WS14 9TZ</p>	<p>Contact: Lauren Cresswell Tel: +44 (0)1543 261020 Fax: +44 (0)1543 261021 E-Mail: lcresswell@mercianscience.co.uk Website: www.mercianscience.co.uk</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATER	<u>Microbiological Tests</u>	Documented In-house Methods
Potable (non regulatory) and Process, Hospital water, Pools and Hydrotherapy pool water	Enumeration: Coliforms / <i>E. coli</i> , presumptive and confirmed	Method B3 using membrane filtration and MLGA (single membrane) based on The Microbiology of Drinking Water Part 4b, 2016
Potable (non regulatory) and Process	Aerobic colony count at 22 °C, 30 °C and 37 °C	Method B1 using pour plate based on The Microbiology of Drinking Water Part 7, 2020
Hospital water, Pools, and Hydrotherapy pool water	Aerobic colony count at 22 °C and 37 °C	Method B1 using pour plate based on The Microbiology of Drinking Water Part 7, 2020
Potable (non regulatory) Process, Pool and Hydrotherapy pool water	<i>Pseudomonas aeruginosa</i> and <i>Pseudomonas</i> spp, presumptive	Method B5 using Membrane filtration with CFC isolation agar and CN agar for confirmation
Endoscopy washer disinfectant rinse water	<i>Pseudomonas aeruginosa</i>	Method B5 using membrane filtration based on HTM 01-06
Potable (non regulatory), Pool and Hydrotherapy pool water	<i>Pseudomonas aeruginosa</i> , confirmed	Method B18 using Pseudalert based on The Microbiology of Drinking Water part 8C 2015
Hospital tap water	<i>Pseudomonas aeruginosa</i> , confirmed	Method B18 based on HTM 04-01 Part B Appendix F using MoDW (2015) part 8C using Pseudalert
Potable (non regulatory), including Process and Pools	Enterococci, presumptive and confirmed	Method B6 using membrane filtration based on The Microbiology of Drinking Water Part 5, 2012



2450
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Mercian Science Ltd
Issue No: 033 **Issue date:** 05 December 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATER (cont'd) Potable, Process, Pools & Hospital water	<u>Microbiological Tests (cont'd)</u> Detection and enumeration: <i>Legionella</i> spp confirmed, and identification of <i>Legionella pneumophila</i> serogroups 1, 2-14	Documented In-house Methods (cont'd) Method B4 using membrane filtration and manual elution
END		