

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>24991</b> Accredited to ISO 15189: 2012	<b>InHealth Pathology Ltd trading as InHealth Pathology</b>	
	<b>Issue No:</b> 001 <b>Issue date:</b> 11 October 2024	
	<b>4th Floor Lambourne House</b> <b>7 Western Road</b> <b>Romford</b> <b>Essex</b> <b>RM1 3LD</b>	<b>Contact:</b> Donna Horncastle <b>Tel:</b> +44 (0) 1708957381 <b>E-Mail:</b> pathology@inhealthgroup.com <b>Website:</b> www.inhealthgroup.com
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<b>HUMAN BODY TISSUES</b>	<u>Histopathological examination activities for the purposes of clinical diagnosis</u>	In-house documented procedures using manual methods or analysers in conjunction with manufacturers' instructions as specified:
Endoscopic Biopsies - Formalin-fixed	Examination and preparation of tissues in order to identify or exclude morphological and cytological abnormalities for the purpose of diagnosis	Reception of Histology Specimens PAT-R-LP-01  Specimen transfer and dissection PAT-R-LP-03
Endoscopic Biopsies -fixed tissue samples		Processing Using: Sakura VIP 6AI processor PAT-R-LP-04
Formalin fixed paraffin wax blocks of processed tissue		Embedding Using: Sakura Tissue-TEK embedding centre PAT-R-LP-05
Formalin fixed paraffin embedded slides		Sectioning and slide production Using: Leica RM2245 microtomes PAT-R-LP-06
Formalin fixed paraffin wax blocks of processed tissue	Routine morphological staining to detect basophilic and eosinophilic tissue structures to assist in the detection of clinical abnormalities	Routine Haematoxylin and Eosin (H&E) staining and coverslipping Using: Roche Ventana HE600 PAT-R-LP-07



24991  
Accredited to  
ISO 15189: 2012

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**InHealth Pathology Ltd trading as InHealth Pathology**

**Issue No: 001 Issue date: 11 October 2024**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY TISSUES (cont'd)	<u>Histopathological examination activities for the purposes of clinical diagnosis</u> (cont'd)	In-house documented procedures using manual methods or analysers in conjunction with manufacturers' instructions as specified:
Formalin fixed paraffin wax blocks of processed tissue	Special staining of histopathology examinations to assist in detection of clinical abnormalities:	Special staining. Manual documented procedures based on manufacturers' kits:
	Mucins, Glycogen, PAS positive substances-fungi	PAS (Periodic Acid Schiffs) PAT-R-LP-31
	Glycogen (Diastase digestion)	PAS Diastase PAT-R-LP-31
Formalin fixed paraffin embedded stained slides and digital images	Morphological assessment and interpretation/diagnosis (Digital Pathology and light microscopy)	Interpretive Microscopy (qualitative analysis) Using: Philips UFS PAT-R-LP-27 (Digital Pathology) Quality control/ Microscopes PAT-R-LP-21
END		