


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2569</b>  Accredited to ISO/IEC 17025:2017	<b>Lucion Services Ltd</b>	
	<b>Issue No:</b> 032 <b>Issue date:</b> 10 April 2025	
	<b>Unit 7</b> Halifax Court Dunston Gateshead Newcastle upon Tyne NE11 9JT	<b>Contact:</b> Chris Parr <b>Tel:</b> +44 (0)191 461 8999 <b>E-Mail:</b> Chris.Parr@luciongroup.com <b>Website:</b> www.luciongroup.com
Testing performed by the Organisation at the locations specified below		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code	
<b>Address</b> Unit 7 Halifax Court Dunston Gateshead Newcastle upon Tyne NE11 9JT	<b>Local contact</b> Chris Parr	Health and Hygiene  Asbestos – All Support Functions  <b>Support functions:</b> Quality System Quality Audit Administration	A
<b>Address:</b> Unit 22 Salbrook Road Industrial Estate Redhill Surrey RH1 5DY	<b>Local contact</b> Chris Parr	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li><li>• <i>Reporting</i></li></ul>	D
<b>Address:</b> The Hall Rugby Road Wolston Coventry CV8 3FZ	<b>Local contact</b> Chris Parr	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li><li>• <i>Reporting</i></li></ul>	E



2569  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lucion Services Ltd**  
**Issue No: 032 Issue date: 10 April 2025**

Testing performed by the Organisation at the locations specified

Location details	Activity	Location code
<b>Address:</b> Unit 9C Hellaby Industrial Estate Rotherham S66 8QY	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li></ul>	G
<b>Address</b> 5 Abbots Park Preston Brook Runcorn WA7 3GH	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li><li>• <i>Reporting</i></li></ul>	L
<b>Address</b> G3 The Bloc Springfield Way Anlaby Yorkshire HU10 6RJ	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li></ul>	N
<b>Address</b> Sellafield Seascale Cumbria CA20 1PG	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Contract Review</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li><li>• <i>Reporting</i></li></ul>	O
<b>Address</b> Omagh Enterprise Centre Great Northern Road Omagh BT78 5LU	Health and Hygiene  Asbestos – Support Functions: <ul style="list-style-type: none"><li>• <i>Quality Audit Administration</i></li><li>• <i>Scheduling</i></li><li>• <i>Personnel</i></li><li>• <i>Equipment</i></li><li>• <i>Measurement Traceability</i></li></ul>	P



2569  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lucion Services Ltd**  
**Issue No: 032 Issue date: 10 April 2025**

Testing performed by the Organisation at the locations specified

**Site activities performed away from the locations listed above:**

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	I
Temporary Facilities	Health and Hygiene	Q *



2569  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lucion Services Ltd**  
**Issue No: 032 Issue date: 10 April 2025**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021  This organisation holds accreditation to establish new temporary site laboratories to conduct the following asbestos testing activities: <ul style="list-style-type: none"> <li>• <i>Fibre counting</i></li> <li>• <i>Identification of asbestos in bulk material</i></li> </ul> The Site laboratories (denoted*) are set up in accordance with the Documented In-House Procedure TOP01.02.07	
	Sampling of air for fibre counting	Documented In-House Method TOP02.02.08 based on HSG 248	A, B, D, E, G, I, L, N, O, P
	Fibre counting	Documented In-House Method TOP02.02.09, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, D, E, G, I, L, N, O, P
	4 Stage Clearance Process	Documented In-House Method TOP02.02.08, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, D, E, G, L, N, O, P
ASBESTOS IN BULK MATERIALS Including materials and products suspected of containing asbestos	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method TOP01.02.03 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A, O



2569  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lucion Services Ltd**  
**Issue No: 032 Issue date: 10 April 2025**

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	<u>Health and Hygiene (cont'd)</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
<p>ASBESTOS IN SOILS – The Identification of Asbestos fibres in bulk samples of Soil, <i>specifically:</i></p> <ul style="list-style-type: none"> <li>• Soil</li> <li>• Sediment</li> <li>• Slurry</li> <li>• Ballast</li> <li>• Aggregate</li> <li>• Crushed concrete</li> <li>• Rail ballast</li> </ul>	<p>Identification of:</p> <p>Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Documented In-House Procedure TOP01.02.03 using stereo-microscopy, polarised light microscopy and dispersion staining based on HSG 248</p>	A
<p>ASBESTOS IN SOILS The Identification and Quantification of Asbestos fibres in bulk samples of Soil, <i>specifically:</i></p> <ul style="list-style-type: none"> <li>• Soil</li> <li>• Sediment</li> <li>• Slurry</li> <li>• Ballast</li> <li>• Aggregate</li> <li>• Crushed concrete</li> <li>• Rail ballast</li> </ul>	<p>Identification and Quantification of Asbestos content of:</p> <p>Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite</p>	<p>Documented In-House Procedure TOP01.02.03 for identification using stereo-microscopy, polarised light microscopy and dispersion staining based on HSG 248</p> <p>Documented In-House Procedure TOP01.02.03 for quantification of asbestos using asbestos fibre identification and fibrecounting following Gravimetric analysis (Stage 2) and Free fibre analysis via sedimentation (Stage 3). Using SEM and EDS.</p>	A



2569  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Lucion Services Ltd**  
**Issue No:** 032    **Issue date:** 10 April 2025

**Testing performed by the Organisation at the locations specified**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
	<u>Health and Hygiene (cont'd)</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
GYPSUM POWDER Only - The Identification of Asbestos fibres	Identification and Quantification of Asbestos Fibres (as chrysotile or amphibole asbestos only)	Documented In-house method TOP02.09.03 based on VDI 3492, ISO14966 and HSG 248 using scanning electron microscopy (SEM) and energy dispersive x-ray spectroscopy (EDS)	A
ASBESTOS FIBRES IN AIR  (Fibres in air (SEM) Asbestos and other fibres derived from asbestos products, dusts, soils and air.)	Sampling of air and dust deposits onto membrane filters for fibre identification (serpentine and amphibole asbestos, inorganic fibres and calcium sulphate fibres) and evaluation of airborne fibre concentration by scanning electron microscopy (SEM) and energy dispersive x-ray spectroscopy (EDS).	In-house Procedure TOP02.02.09.03 based on VDI 3492, ISO14966 and HSG248 using SEM and EDS	A, B
	Identification of asbestos fibres (as serpentine or amphibole asbestos), inorganic fibres and calcium sulphate fibres on membrane filters by scanning electron microscopy (SEM) and energy dispersive x-ray spectroscopy (EDS).	In-house Procedure TOP02.02.09.03 based on VDI 3492, ISO14966 and HSG248 using SEM and EDS	A
	Fibrecounting for the evaluation of airborne fibre concentrations by scanning electron microscopy (SEM) and energy dispersive x-ray spectroscopy (EDS).	In-house Procedure TOP02.02.09.03 based on VDI 3492, ISO14966 and HSG248 using SEM and EDS	A
END			