


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> MEDICAL <b>25821</b></p> <p>Accredited to ISO 15189: 2022</p>	<b>Harley Street Pathology Services Limited</b>	
	<b>Issue No: 004    Issue date: 10 June 2025</b>	
	<b>134-136 Harley Street</b> <b>London</b> <b>W1G 7JY</b>	<b>Contact: Sonal Parikh</b> <b>Tel: +44 (0) 77 34705973</b> <b>E-Mail: Sonal@hspathology.com</b> <b>Website: <a href="https://hspathology.com/">https://hspathology.com/</a></b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<b>HUMAN BODY FLUIDS</b>	<u>Biochemistry examination activities for the purposes of clinical diagnosis</u>	Manufacturer's instructions and documented In-house methods using:
Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum	Alanine Aminotransferase (ALT) Albumin Alkaline Phosphatase (ALP) Aspartate Aminotransferase (AST) Bilirubin Calcium Cholesterol C-reactive protein Creatinine Gamma Glutamyl transferase (GGT)	Roche Cobas Pure Analyser HS BIO SOP-01  Roche Cobas Pure c302 Photometric chemistry methods:  HS BIO SOP-5 HS BIO SOP-2 HS BIO SOP-4 HS BIO SOP-21 HS BIO SOP-22 HS BIO SOP-23 HS BIO SOP-24 HS BIO SOP-26 HS BIO SOP-25 HS BIO SOP-27
Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum	Glucose High Density Lipoproteins (HDL) Immunoglobulin A (IgA) Immunoglobulin G (IgG) Immunoglobulin M (IgM) Iron Magnesium Phosphate Total protein Transferrin Triglycerides Unsaturated Iron-binding Capacity Urea	HS BIO SOP-29 HS BIO SOP-30 HS BIO SOP-31 HS BIO SOP-32 HS BIO SOP-33 HS BIO SOP-34 HS BIO SOP-36 HS BIO SOP-37 HS BIO SOP-39 HS BIO SOP-40 HS BIO SOP-41 HS BIO SOP-42 HS BIO SOP-43



25821  
Accredited to  
ISO 15189: 2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Harley Street Pathology Services Limited**

**Issue No: 004    Issue date: 10 June 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Biochemistry examination activities for the purposes of clinical diagnosis</u> (cont'd)	Manufacturer's instructions and documented In-house methods using:  Roche Cobas Pure c303 Indirect Ion Selective Electrode methods  HS BIO SOP-35 HS BIO SOP-35 HS BIO SOP-35  Roche Cobas Pure e402 Electrochemiluminescence (ECL)
Serum Serum Serum	Chloride Potassium Sodium	HS BIO SOP-20 HS BIO SOP-7 HS BIO SOP-8 HS BIO SOP-9 HS BIO SOP-11 HS BIO SOP-12 HS BIO SOP-10 HS BIO SOP-13 HS BIO SOP-14 HS BIO SOP-16 HS BIO SOP-15 HS BIO SOP-17 HS BIO SOP-18 HS BIO SOP-6 HS BIO SOP-19
Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum Serum	Anti-Mullerian Hormone (AMH) Oestradiol Ferritin Folate Free T3 (FT3) Free T4 (FT4) Follicle Stimulating Hormone (FSH) $\beta$ HCG Luteinising Hormone (LH) Progesterone Prolactin Testosterone Thyroid Stimulating Hormone (TSH) Vitamin B12 Vitamin D	
EDTA blood	<u>Haematology examination activities for the purposes of clinical diagnosis</u>  Full blood count	Manufacturer's instructions and documented In-house methods using:  Sysmex XN-450 Analyser: DC detection, SLS method, flow cytometry & fluorescence flow cytometry HS HAEM SOP-01 HS HAEM SOP-2



25821  
Accredited to  
ISO 15189: 2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Harley Street Pathology Services Limited**

**Issue No: 004    Issue date: 10 June 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p><b>HUMAN BODY FLUIDS (cont'd)</b></p> <p>High Vaginal Swabs (HVS) Semen Urine</p>	<p><u>Microbiology examination activities for the purposes of clinical diagnosis</u></p> <p>General isolation and characterisation of micro-organisms of clinical significance</p>	<p>Manufacturer's instructions and documented In-house methods using:</p> <p>Manual culture and Identification of:</p> <ul style="list-style-type: none"> <li>Gram-negative, oxidase-negative bacilli using selective, non-selective and chromogenic culture media, RapID system (Thermo Scientific) and</li> <li>Beta-haemolytic streptococci groups A-G using ProLab Streptococcus grouping kit routine biochemical and latex antigen detection reagents</li> </ul> <p>Manual culture of Candida spp. HS MIC SOP-7 HS MIC SOP-8</p>
<p>Bacterial isolates as isolated above on solid culture medium from HVS, semen and urine samples</p>	<p>Antimicrobial sensitivity testing</p>	<p>EUCAST manual disc sensitivity testing HS MIC SOP-14</p>
<p>HVS</p>	<p>Examination for Bacterial Vaginosis</p>	<p>Light microscopy of Gram-stained slides using Hays criteria HS MIC-SOP-7</p>
<p>Urine</p>	<p>Quantification of white cells, red cells and bacteria</p>	<p>Manual light microscopy HS MIC SOP-8</p>



25821  
Accredited to  
ISO 15189: 2022

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Harley Street Pathology Services Limited**

**Issue No: 004    Issue date: 10 June 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS (cont'd)	<u>Virology examination activities for the purposes of clinical diagnosis</u>	Manufacturer's instructions and documented In-house methods using:
	<u>Serology:</u>	Roche Cobas Pure Analyser: Electro-chemiluminescence (ECL) measuring system. HS BIO SOP-1
Serum	Anti-HBc	HS SER VIR SOP-13
Serum	CMV IgG	HS SER VIR SOP-4
Serum	HBsAg	HS SER VIR SOP-14
Serum	HIV-1 Ab	HS SER VIR SOP-11
Serum	HIV-2 Ab	HS SER VIR SOP-11
Serum	HIV Ag (p24)	HS SER VIR SOP-11
Serum	Anti-HCV	HS SER VIR SOP-12
Serum	Syphilis IgG, IgM	HS SER VIR SOP-8
Serum	Toxoplasma IgM	HS SER VIR SOP-9
Serum	Rubella IgG	HS SER VIR SOP-6
Serum	CMV IgM	HS SER VIR SOP-2
Serum	HTLV I and II	HS SER VIR SOP-5
	Molecular virology:	Real time PCR using the Hologic Panther
Urine	Detection of <i>Chlamydia trachomatis</i> (CT) and <i>Neisseria gonorrhoea</i> (NG) RNA	HS MOL SOP-01 using the Hologic Aptima Combo II kit
Thin prep liquid-based cytology specimens	Detection of HPV mRNA	HS MOL SOP-5 using the Hologic Aptima HPV Assay kit
END		