


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2621 Accredited to ISO/IEC 17025:2017	Spectra Analysis Services Ltd Issue No: 015 Issue date: 16 April 2026	
	Unit 2 Olympic Business Centre Paycocke Road Basildon Essex SS14 3EX	Contact: Mr Darren Chinnery Tel: +44 (0)1268 534380 Fax: +44 (0)1268 272985 E-Mail: darrenchinnery@spectraholdings.co.uk Website: www.spectraholdings.co.uk

Testing performed by the Organisation at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 2 Olympic Business Centre Paycocke Road Basildon Essex SS14 3EX	Local contact Mr Darren Chinnery Tel: +44 (0)1268 534380 Fax: +44 (0)1268 272985 E-Mail: darrenchinnery@spectraholdings.co.uk	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details		Activity	Location code
Client Premises	Contacts as above	Health and Hygiene	B
Mobile Testing Laboratories	Contacts as above	Health and Hygiene Asbestos (sampling and testing)	C



2621
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Spectra Analysis Services Ltd
Issue No: 015 Issue date: 16 April 2026

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	A, B, C
	Sampling of air for fibre counting	Documented In-House Method No 1 based on HSG 248	
	Fibre counting	Documented In-House Method No 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method No 1, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B, C
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method No 2, using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A
END			