


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>26223 Accredited to ISO/IEC 17025:2017</p>	Storm Aviation Limited Issue No: 005 Issue date: 09 February 2026	
	Unit 1 Bluegates Business Park Stortford Road Little Canfield CM6 1SN	Contact: Michael Reed Tel: +44 (0)1279 669014 E-Mail: calibration@stormaviation.com Website: www.stormaviation.com
Calibration performed at the above address only		

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks
PRESSURE Hydraulic pressure (gauge) Gas pressure (gauge) Gas pressure (absolute)	344 kPa to 138 MPa -95 kPa to 100 kPa 100 kPa to 2 MPa 2 MPa to 3.5 MPa 3.5 MPa to 7 MPa 5 kPa to 200 kPa 200 kPa to 2.1 MPa 2.1 MPa to 3.6 MPa 3.6 MPa to 7.1 MPa	Q [0.012 %, 345 Pa] 30 Pa 200 Pa 400 Pa 700 Pa 30 Pa 200 Pa 400 Pa 700 Pa	Methods consistent with EURAMET CG17 Calibration of pressure indicating instruments and gauges
TORQUE Torque Wrenches	0.4 N·m to 2 N·m 2 N·m to 50 N·m 10 N·m to 250 N·m 40 N·m to 1000 N·m	1.1 % 1.1 % 0.31 % 1.0 %	BS EN ISO 6789-2:2017 1. The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated 2. Calibrations may be given in other torque units lbf.in, lbf.ft, oz.in.
END			



26223

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Storm Aviation Limited

Issue No: 005 Issue date: 09 February 2026

Calibration performed at main address only

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$