


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 26393 Accredited to ISO/IEC 17025:2017	Denas Management Solutions Limited trading as Denas Analytical Issue No: 002 Issue date: 17 April 2023	
	The Lab First Floor Unit B, Spenlea Industrial Estate Wharf Road Ealand DN17 4JW	Contact: Mr J Collett Tel: +44 (0) 1724 514488 E-Mail: lab@denas.co.uk
Testing performed at the above address only		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	<u>Health and Hygiene</u> Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Health and Safety Executive Asbestos: The Analysts' Guide (HSG 248) - 2021 Documented In-House Procedure SOP09 using stereo-microscopy, polarised light microscopy and dispersion staining
END		