


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>2650</b> Accredited to <b>ISO/IEC 17025:2017</b>	<b>National Gas Transmission Plc</b>  Issue No: 015    Issue date: 28 April 2023	
	<b>National Grid House</b> Warwick Technology Park Gallows Hill Warwick CV34 6DA	<b>Contact: Mr C Pickford</b> Tel: +44 (0) 7966 836392 Email: christopher.pickford@nationalgas.com Website: www.nationalgas.com

Testing performed by the Organisation at the locations specified below

### Locations covered by the organisation and their relevant activities

#### Laboratory location:

Location details	Activity	Location code
<b>Address</b> National Grid House Warwick Technology Park Gallows hill Warwick CV34 6DA  <b>Local contact</b> Mr Christopher Pickford Tel: +44 (0) 7966 836392 Email: christopher.pickford@nationalgas.com Website: www.nationalgas.com	<b>Support Functions:</b> Quality System Quality Audit Administration  <b>Sampling and Testing:</b> Stack emissions Testing	A

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer sites requiring Stack Emissions Testing	Stack emissions Testing	B



2650  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**National Grid Plc**  
**Issue No: 015 Issue date: 28 April 2023**

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
Testing of Stack Emissions to Atmosphere	<u>Sampling and On-Line Analysis</u>	National, European, International and Environment Agency specified standards including MIDs and Documented In-House work instructions to meet the requirements of the Environment Agency (MCERTS) Performance Standard and BS EN 15259:2007	
	Carbon monoxide*	BS EN 15058:2017 (National Grid Emissions Monitoring Procedure - NDIR analyser)	B
	Oxygen*	BS EN 14789:2017 (National Grid Emissions Monitoring Procedure - Paramagnetic analyser)	B
	Oxides of Nitrogen *	BS EN 14792:2017 (National Grid Emissions Monitoring Procedure - Chemiluminescence analyser & Validated Ultra Violet analyser)	B
Stack Emissions - Continuous Emissions Monitoring Systems (CEMS)	QAL 2 and the Annual Surveillance Test (AST) for CEMS	Documented in house National Grid Emissions Monitoring Procedure to meet the requirements of BS EN 14181:2014, Environment Agency MID 14181 (TGN M20 Annex A) and other requirements of the Environment Agency (MCERTS) Performance Standard for laboratories carrying out testing of samples from stack emissions monitoring and BS EN 15259:2007	A + B
END			

\* - The scale range of the analyser used for this test must be that detailed on its current MCERTS certificate or a range validated by the organisation to meet MCERTS requirements.