


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 27146 Accredited to ISO/IEC 17025:2017	Premier Calibrations Limited	
	Issue No: 001	Issue date: 20 November 2023
	219 Bentley Road Bentley Doncaster DN5 9TB	Contact: Mr Adrian King Tel: +44 (0)1302 787333 E-Mail: premc@premc.co.uk Website: www.premier-calibration.co.uk
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 219 Bentley Road Bentley Doncaster DN5 9TB Local contact Mr Adrian King Tel: +44 (0)1302 787333 Email: premc@premc.co.uk Website: www.premier-calibration.co.uk	Pressure instrument calibration	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Calibrations can be performed in a temperature controlled mobile calibration laboratory.	Torque wrench, torque screwdriver and pressure instrument calibration	Mobile

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Pressure instrument calibration	Site



27146
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Premier Calibrations Limited
Issue No: 001 **Issue date:** 20 November 2023

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ($k = 2$)	Remarks	Location Code
PRESSURE Hydraulic Pressure (Gauge) Calibration of pressure indicating instruments and gauges Gas Pressure (Gauge) Calibration of pressure indicating instruments and gauges	0 MPa to 95 MPa - 83 kPa to 0 kPa 0 kPa to 2 MPa	54 kPa 1.3 kPa 1.4 kPa	Methods consistent with EURAMET CG17.	Lab, Mobile & Site
TORQUE Hand Torque Tools Wrenches Wrenches and Screwdrivers	BS EN ISO 6789-2:2017 0.15 N·m to 1500 N·m BS EN ISO 6789:2003 (withdrawn) 0.15 N·m to 1500 N·m	1.0 % 1.6 %	The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated. Calibrations may also be given in lbf-in and lbf-ft.	Mobile
END				



27146
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Premier Calibrations Limited
Issue No: 001 **Issue date:** 20 November 2023

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$