

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2742 Accredited to ISO/IEC 17025:2017	AC & MS Ltd	
	Issue No: 016	Issue date: 03 November 2023
	Unit 13 International Business Park Charfleets Road Canvey Island SS8 0SG	Contact: Mr R Walker Tel: +44 (0)1268 680136 E-Mail: roy.walker@acandms.co.uk Website: www.acandms.co.uk
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address: Unit 13 International Business Park Charfleets Road Canvey Island SS8 0SG Local contact Contact: Mr R Walker Tel: +44 (0)1268 680136	Health and Hygiene Head Office Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile testing laboratories	Health and Hygiene	C



2742
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

AC & MS Ltd

Issue No: 016 **Issue date:** 03 November 2023

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method PROC 6 based on HSG 248	A, B, C
	Fibre counting	Documented In-House Method PROC 6, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method PROC 6, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	B
	Sampling of bulk materials for asbestos identification	Documented In-House Procedure 001 based on HSG 248	B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method PROC 005 using stereo-microscopy, polarised light microscopy and dispersion staining based on HSG 248	A
END			