

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



2770

Accredited to
ISO/IEC 17025:2017

INEOS Chemicals Grangemouth Limited

Issue No: 026 Issue date: 04 November 2025

PO Box 21
Bo'ness Road
Grangemouth
Stirlingshire
FK3 9XH

Contact: Mrs G Davis
Tel: +44 (0)1324 493134
E-Mail: gayle.davis@ineos.com
Website: www.ineos.com/grangemouth

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
FINAL EFFLUENT	<u>Chemical and Physical Tests</u> pH Total Suspended Solids Chemical Oxygen Demand Oil in Water Hydrocarbon Oil Index (HOI)	Documented In-House Method LM-Water-29 to BS EN 10523:2012 Documented In-House Method LM-HSE-10 to BS EN 872:2005 Documented In-House Method LM-HSE-28 based on ISO 6060:1989 Documented In-House Method LM-HSE-26 based on SCA Blue Book No. 77 (ISBN 0117517275) Documented In-House Method LM-HSE-36 based on BS EN ISO 9377-2 2000 By GC-FID
LIQUIFIED PETROLEUM GAS	Propane and butane	BS EN 27941 - 1994 (Modified) supported by Documented In-House Method LM-GC-03
NATURAL GAS	C ₁ - C ₅ alkanes, > C ₅ alkanes, air (oxygen not quantified), CO ₂	ASTM D 1945-19 (Modified) supported by Documented In-House Method LM-GC-01



2770

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

INEOS Chemicals Grangemouth Ltd
Issue No: 026 Issue date: 04 November 2025

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used																																																				
PETROLEUM and PETROLEUM PRODUCTS	<u>Chemical and Physical Tests</u> (cont'd)																																																					
	Density and relative density of liquids by digital density meter	IP 365/97(20) EN ISO 12185:24																																																				
REFINERY GAS	Determination of Composition of Fuel Gas Streams Chemical composition: Amount fraction (% mol/mol and % m/m)	Documented in-house method LM-GC-11 using gas chromatography																																																				
	<table><tr><th>Component</th><th>(% mol)</th></tr><tr><td>Methane</td><td>0.01 to 99.0</td></tr><tr><td>Ethane</td><td>0.01 to 99.9</td></tr><tr><td>Ethene</td><td>0.02 to 99.9</td></tr><tr><td>Propane</td><td>0.01 to 99.9</td></tr><tr><td>Propene</td><td>0.01 to 99.5</td></tr><tr><td>n-Butane</td><td>0.01 to 99.9</td></tr><tr><td>iso-Butane</td><td>0.01 to 99.0</td></tr><tr><td>trans-2-Butene</td><td>0.01 to 5.0</td></tr><tr><td>iso-Butene</td><td>0.01 to 5.0</td></tr><tr><td>1-Butene</td><td>0.01 to 5.0</td></tr><tr><td>cis-2-Butene</td><td>0.01 to 5.0</td></tr><tr><td>1,3 Butadiene</td><td>0.01 to 10.0</td></tr><tr><td>n-Pentane</td><td>0.01 to 4.5</td></tr><tr><td>iso-Pentane</td><td>0.01 to 4.5</td></tr><tr><td>Cyclopropane</td><td>0.01 to 5.0</td></tr><tr><td>Propadiene</td><td>0.01 to 5.0</td></tr><tr><td>Acetylene</td><td>0.01 to 2.0</td></tr><tr><td>C6⁺(>n-Pentane)</td><td>0.01 to 4.0</td></tr><tr><td>Hydrogen Sulphide</td><td>0.1 to 2.0</td></tr><tr><td>Hydrogen</td><td>0.01 to 99.9</td></tr><tr><td>Helium</td><td>0.01 to 99.9</td></tr><tr><td>Carbon Dioxide</td><td>0.02 to 50.0</td></tr><tr><td>Carbon Monoxide</td><td>0.02 to 50.0</td></tr><tr><td>Nitrogen</td><td>0.02 to 75</td></tr><tr><td>Oxygen</td><td>0.02 to 20.0</td></tr></table>	Component	(% mol)	Methane	0.01 to 99.0	Ethane	0.01 to 99.9	Ethene	0.02 to 99.9	Propane	0.01 to 99.9	Propene	0.01 to 99.5	n-Butane	0.01 to 99.9	iso-Butane	0.01 to 99.0	trans-2-Butene	0.01 to 5.0	iso-Butene	0.01 to 5.0	1-Butene	0.01 to 5.0	cis-2-Butene	0.01 to 5.0	1,3 Butadiene	0.01 to 10.0	n-Pentane	0.01 to 4.5	iso-Pentane	0.01 to 4.5	Cyclopropane	0.01 to 5.0	Propadiene	0.01 to 5.0	Acetylene	0.01 to 2.0	C6 ⁺ (>n-Pentane)	0.01 to 4.0	Hydrogen Sulphide	0.1 to 2.0	Hydrogen	0.01 to 99.9	Helium	0.01 to 99.9	Carbon Dioxide	0.02 to 50.0	Carbon Monoxide	0.02 to 50.0	Nitrogen	0.02 to 75	Oxygen	0.02 to 20.0	
Component	(% mol)																																																					
Methane	0.01 to 99.0																																																					
Ethane	0.01 to 99.9																																																					
Ethene	0.02 to 99.9																																																					
Propane	0.01 to 99.9																																																					
Propene	0.01 to 99.5																																																					
n-Butane	0.01 to 99.9																																																					
iso-Butane	0.01 to 99.0																																																					
trans-2-Butene	0.01 to 5.0																																																					
iso-Butene	0.01 to 5.0																																																					
1-Butene	0.01 to 5.0																																																					
cis-2-Butene	0.01 to 5.0																																																					
1,3 Butadiene	0.01 to 10.0																																																					
n-Pentane	0.01 to 4.5																																																					
iso-Pentane	0.01 to 4.5																																																					
Cyclopropane	0.01 to 5.0																																																					
Propadiene	0.01 to 5.0																																																					
Acetylene	0.01 to 2.0																																																					
C6 ⁺ (>n-Pentane)	0.01 to 4.0																																																					
Hydrogen Sulphide	0.1 to 2.0																																																					
Hydrogen	0.01 to 99.9																																																					
Helium	0.01 to 99.9																																																					
Carbon Dioxide	0.02 to 50.0																																																					
Carbon Monoxide	0.02 to 50.0																																																					
Nitrogen	0.02 to 75																																																					
Oxygen	0.02 to 20.0																																																					
END																																																						