


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines upon Thames, TW18 3HR, UK

 <p>UKAS TESTING</p> <p>2790</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Esso Petroleum Company Ltd</h3> <p>Issue No: 009 Issue date: 08 October 2019</p>	
	<p>Fawley Refinery Cadland Road Hythe Southampton SO45 3NP</p>	<p>Contact: No commercial enquiries</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>PETROLEUM and PETROLEUM PRODUCTS</p> <p>REFINERY GAS</p>	<p><u>Chemical Tests</u></p> <p>Composition as mol%:</p> <p>Carbon monoxide 0.01 to 100 Carbon dioxide 0.01 to 100 Hydrogen 0.05 to 100 Nitrogen 0.01 to 100 Oxygen 0.01 to 100 Methane 0.01 to 100 Ethane 0.01 to 100 Ethylene 0.01 to 100 Propane 0.01 to 100 Propylene 0.01 to 100 Propadiene 0.01 to 100 Acetylene 0.01 to 100 Isobutane 0.01 to 100 n-Butane 0.01 to 100 Methyl cyclopropane 0.01 to 100 1-Butene 0.01 to 100 Isobutene 0.01 to 100 Trans-2-butene 0.01 to 100 Cis-2-butene 0.01 to 100 1,2-Butadiene 0.01 to 100 1,3-Butadiene 0.01 to 100 Propyne 0.01 to 100 1-Butyne 0.01 to 100 Saturated C5s 0.01 to 100 Unsaturated C5s 0.01 to 100 C6 and heavier 0.01 to 100</p>	<p>Documented In-House Method Chemicals-Lab-GC Test Method 40 using gas chromatography</p>
<p>END</p>		