

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 2794 Accredited to ISO/IEC 17025:2017	ENV Surveys Ltd	
	Issue No: 007 Issue date: 01 February 2022	
	Formula House 12 Upper Hollingdean Road Brighton East Sussex BN1 7GA	Contact: Mr J Cameron Tel: +44 (0)1273 506098 E-Mail: info@envsurveys.com Website: www.envsurveys.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Formula House 12 Upper Hollingdean Road Brighton East Sussex BN1 7GA	Local contact Mr J Cameron Tel: +44 (0)1273 506098 Email: info@envsurveys.com	Health and Hygiene Head Office Asbestos – All Support Functions A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile testing laboratory	Health and Hygiene	C



2794
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

ENV Surveys Ltd
Issue No: 007 Issue date: 01 February 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method Air Testing & Fibre Counting Methodology, Section 18 based on HSG 248	A, B
	Fibre counting	Documented In-House Method, Air Testing & Fibre Counting Methodology, Section 18, using Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
ASBESTOS IN BULK MATERIALS including materials and products suspected of containing asbestos	4 Stage Clearance Process	Documented In-House Method, Site Assessment for Reoccupation, Section 19 using Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B
	Identification of: Amosite Chrysotile Crocidolite Fibrous Actinolite Fibrous Anthophyllite Fibrous Tremolite	Documented In-House Method Bulk Identification, Section 16 using stereo-microscopy, polarised light optical microscopy and dispersion staining based on HSG 248	A

END