


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION <b>27982</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>QROMETRIC LIMITED</b></p> <p>Issue No: 004 Issue date: 22 June 2026</p>	
	<p>Unit 10 Hove Enterprise Centre Basin Road North Portslade BN41 1UY</p>	<p>Contact: Callum Morgan Tel: +44 (0) 207 099 5807 E-Mail: callum@qrometric.com info@qrometric.com Website: www.qrometric.com</p>

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>HUMIDITY</b>			Items with an electrical output can also be calibrated.
<b>Dew/Frost Point</b>			
Measurement devices	-90 °C to -75 °C -75 °C to -60 °C -60 °C to -20 °C -20 °C to +75 °C +75 °C to +85 °C	0.66 °C to 0.21 °C 0.21 °C to 0.069 °C 0.069 °C 0.057 °C 0.065 °C	Calibration by comparison with a reference condensation hygrometer.  Measurement uncertainty may vary depending on the specification and performance of the item calibrated.
Generators	-90 °C to -75 °C -75 °C to -60 °C -60 °C to -20 °C -20 °C to +75 °C +75 °C to +85 °C	0.66 °C to 0.21 °C 0.21 °C to 0.069 °C 0.069 °C 0.057 °C 0.065 °C	
<b>Relative Humidity</b>	Example conditions corresponding to the dew point and temperature in air ranges		
Condensation hygrometers	-10 °C to 20 °C 5 %rh to 98 %rh  20 °C to 85 °C 5 %rh to 98 %rh	0.073 °C 0.12 %rh to 0.69 %rh  0.073 °C 0.12 %rh to 0.54 %rh	Calibration by comparison with a reference condensation hygrometer and platinum resistance thermometers.  Measurement uncertainty may vary depending on the specification and performance of the item calibrated.
Humidity probes/loggers	-10 °C to 20 °C 5 %rh to 98 %rh  20 °C to 85 °C 5 %rh to 98 %rh	0.073 °C 0.12 %rh to 0.69 %rh  0.073 °C 0.12 %rh to 0.54 %rh	



27982

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**QROMETRIC LIMITED**  
Issue No: 004 Issue date: 22 June 2026

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>TEMPERATURE</b>			Measurement of devices with analogue outputs may be undertaken.
Platinum resistance thermometers and temperature sensors with digital displays, including data loggers	-10 °C to +85 °C	0.073 °C	Calibration by comparison with reference platinum resistance thermometers performed in air using an environmental chamber.
Platinum resistance thermometers and temperature sensors with digital displays, including data loggers	-100 °C to +155 °C	0.033 °C	Using a dry block calibrator. Calibration can be undertaken on PRTs with sheath widths of 2 mm to 3.2 mm or 1/8 inch only.
END			



27982

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**QROMETRIC LIMITED**

**Issue No: 004 Issue date: 22 June 2026**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand.

Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$