

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



2799

Accredited to
ISO/IEC 17025:2017

NDPH Wolfson Laboratories, University of Oxford

Issue No: 016 Issue date: 07 March 2025

Richard Doll Building (RDB)
University of Oxford
Old Road Campus
Roosevelt Drive
Oxford
OX3 7LF

Contact: Professor Michael Hill
Tel: +44 (0)1865 743827
Fax: +44 (0)1865 743985
E-Mail: Michael.Hill@ctsuo.ox.ac.uk
Website: www.ctsu.ox.ac.uk

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
HUMAN BODY FLUIDS	<u>Biochemistry Tests</u>	Documented In House Methods using analysers specified
Plasma/serum	Alanine Aminotransferase (ALT) Alkaline Phosphatase (ALK) Apolipoprotein A ₁ (APOA) Apolipoprotein B (APOB) Aspartate Transaminase (AST) Cholesterol (CHOL) Cholesterol (HDL) Cholesterol (LDL) Creatinine (Jaffe) Total Bilirubin Triglycerides (TG)	Beckman Coulter AU680 Chemistry Analyser Spectrophotometric measurement – Rate, end-point, fixed point
Whole Blood	HbA1c	Documented In House Method using the ADAMS HA 8180 automated Glycohemoglobin (HbA1c) (High Performance Liquid Chromatography)
Serum	Aldosterone	DiaSorin Liaison XS
END		