

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS MEDICAL 28655 Accredited to ISO 15189: 2012	East Suffolk and North Essex NHS Foundation Trust	
	Issue No: 002 Issue date: 06 November 2024	
	Ipswich Hospital Heath Road Ipswich Suffolk IP4 5PD	Contact: Lee Tiltman Tel: +44 (0) 1473 704820 E-Mail: lee.tiltman@esneft.nhs.uk Website: https://www.esneft.nhs.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Department of Microbiology Ipswich Hospital Heath Road Ipswich Suffolk IP4 5PD	Local contact Lee Tiltman Molecular Serology	A



28655
Accredited to
ISO 15189: 2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

East Suffolk and North Essex NHS Foundation Trust

Issue No: 002 Issue date: 06 November 2024

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES and FLUIDS	Molecular examinations for the purpose of clinical diagnosis	In-house documented methods following manufacturers' instructions	
Urethral Swab, Cervical Swab & Urine	Chlamydia trachomatis Neisseria gonorrhoeae	Roche Cobas 6800 Analyser Real-time polymerase chain reaction (PCR) MB-IPS-LP-24	A
Nose / Throat swab in viral transport media	Enterovirus Influenza A detection Influenza A typing Influenza B Human Adenovirus Human Bocavirus Human metapneumovirus Human Parechovirus Human Parainfluenza virus (1-4) Respiratory Syncytial Virus Rhinovirus SARS-CoV-2 Seasonal Coronaviruses	Aus Diagnostics Real-time polymerase chain reaction (PCR) MB-IPS-LP-17	A
Nose / Throat swab in viral transport media	SARS-CoV-2 Antigen Assay (nucleocapsid protein)	LumiraDx Microfluidic immunofluorescence assay MB-IPS-LP-20	A
Nose / Throat swab in viral transport media	SARS-CoV-2 PCR, Influenza A&B, RSV Multiplex Assay (N2 and E nucleic acid targets)	Cepheid GeneXpert Real-time polymerase chain reaction (PCR) MB-IPS-LP-11	A
	SARS-CoV-2 nucleic acid PCR Test SinglePlex Assay	Roche Cobas 6800 Analyser Real-time polymerase chain reaction (PCR) MB-IPS-LP-7	A



28655
Accredited to
ISO 15189: 2012

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

East Suffolk and North Essex NHS Foundation Trust

Issue No: 002 Issue date: 06 November 2024

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
HUMAN TISSUES and FLUIDS (cont'd)	<u>Serological examinations for investigation for immunity and markers of infectious disease</u>	Documented in-house methods to meet the requirements of the Infectious Diseases in Pregnancy (IDPS) screening programme(s) as defined in the Laboratory QA evidence requirements (relevant tests marked with *)	
Serum (unless otherwise specified)	HIV 1 & 2 Antibody p24 antigen assay*	Diasorin Liaison XL Chemiluminescence assay MIC-IPS-LP-18	A
	Hepatitis B antibody* Hepatitis B Core Antibody Hepatitis B Surface antigen	MIC-IPS-LP-4	A
	Hepatitis C antibody	MIC-IPS-LP-20	A
	Syphilis total antibody*	MIC-IPS-LP-8	A
Serum (unless otherwise specified)	HIV 1 & 2 Antibody p24 antigen assay*	BioMerieux VIDAS 3 Enzyme Linked Fluorescent Assay MIC-IPS-LP-18	A
	Hepatitis B Surface antigen* Hepatitis B Core IgM Hepatitis B `e` antibody* Hepatitis B `e` antigen*	MIC-IPS-LP-4	A
	Hepatitis E IgG & IgM	MIC-IPS-LP-5	A
	QuantiFERON Gold Plus	Qiagen / Diasorin Liaison XL Chemiluminescence assay MIC-ALL-LP-2	A
	Syphilis Antibody	Biokit RPR Latex Agglutination assay MIC-IPS-LP-8	A
END			