

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION  <b>29141</b>  Accredited to ISO/IEC 17025:2017	<b>TechniVal Ltd</b>	
	<b>Issue No:</b> 002 <b>Issue date:</b> 11 February 2025	
	<b>Block 4 Unit</b> 7 Chapelhall Industrial Estate Airdrie Lanarkshire ML6 8QH	<b>Contact:</b> Mr David Nelson <b>Tel:</b> +44 (0)8450034247 <b>E-Mail:</b> info@technival.co.uk <b>Website:</b> www.technival.co.uk
<b>Calibration performed by the Organisation at the locations specified</b>		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address:</b> Block 4 Unit 7 Chapelhall Industrial Estate Airdrie, Lanarkshire ML6 8QH  <b>Local contact:</b> David Nelson	Temperature (indicators, Dry Blocks, chambers)	P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
The location must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Temperature (chambers, indicators)	S



29141

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**TechniVal Ltd**

**Issue No: 002    Issue date: 11 February 2025**

**Calibration performed by the Organisation at the locations specified**

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>TEMPERATURE</b>				
Temperature indicators and recorders, with temperature sensor(s)	-196 °C	0.067 °C	LN <sup>2</sup> comparator Stirred liquid baths <b>UKF001 Service</b> <b>UKF008 Service (multi-probe loggers)</b>	P
	-80 °C to 0 °C	0.033 °C		P
	0 to +50 °C	0.029 °C		
	50 °C to 250 °C	0.032 °C		
	250 °C to 300 °C	0.085 °C		
	0 °C to 60 °C	0.30 °C	In Air	P and S
	-90 °C to 0°C	0.10 °C	Dry Well <b>UKF006 Service</b>	P and S
	0 to +125 °C	0.09 °C		
	125 °C to 300 °C	0.11 °C		
Temperature controlled chambers, fridges, freezers, ovens, furnaces and liquid baths (inclusive of associated indicators, controllers and recorders, all with sensors, within the specified parameters and ranges)	-196 °C	0.085 °C	Single monitoring probe. Time dependent by comparison <b>UK005 Service</b>	S
	-80 °C to 150°C	0.064 °C		
	150 °C to 300 °C	0.099 °C		
	-196 °C to 0 °C	0.40 °C	Multipoint monitoring probes. Time dependent temperature profiling, also referred to as spatial temperature surveying or mapping <b>UK007 Service</b>	S
	0 °C to 250 °C	0.37 °C		
	250 °C to 300 °C	0.42 °C		
Dry Block Calibrators	-90 °C to 0 °C	0.032 °C	Method consistent with Euramet CG13 <b>UK004 Service</b>	P
	0 °C to +250 °C	0.028 °C		
	250 °C to 400 °C	0.042 °C		
END				



29141

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

TechniVal Ltd

Issue No: 002 Issue date: 11 February 2025

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$