


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 29622 Accredited to ISO/IEC 17025: 2017	Logical Compliance Limited	
	Issue No: 001 Issue date: 06 March 2025	
	Unit 20A, Beehive Lane Works Beehive Lane Chelmsford Essex CM2 9JY	Contact: Mr John Meates Tel: +44 (0) 7899 907 890 E-Mail: john.meates@logicalcompliance.co.uk Website: www.logicalcompliance.co.uk
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Unit 20A, Beehive Lane Works Beehive Lane Chelmsford Essex CM2 9JY	Local contact Mr John Meates Tel: +44 (0) 7899 907 890	Health and Hygiene Asbestos – All Support Functions	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Client Premises	Health and Hygiene	B
Mobile Laboratories	Health and Hygiene	C



29622
Accredited to
ISO/IEC 17025: 2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Logical Compliance Limited
Issue No: 001 **Issue date:** 06 March 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ASBESTOS FIBRES IN AIR	<u>Health and Hygiene</u>	Health and Safety Executive - Asbestos: The Analysts' Guide (HSG 248) – 2021	
	Sampling of air for fibre counting	Documented In-House Method SOP 01 based on HSG 248	A, B
	Fibre counting	Documented In-House Method SOP 01 Air Monitoring, Membrane Filter Method using Phase Contrast Microscopy (PCM) based on HSG 248	A, B, C
	4 Stage Clearance Process	Documented In-House Method SOP 01 based on HSG 248	B
END			