## **Schedule of Accreditation**

issued by

## **United Kingdom Accreditation Service**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



30011

Accredited to ISO/IEC 17025:2017

## **IRAS Active Analytics Limited**

Issue No: 001 Issue date: 02 June 2025

**Bold Business Centre** 

Bold Lane St Helens Merseyside

WA9 4TX

Contact: Andrew Frith Tel: +44 (0) 1925321662

E-Mail: andrew.frith@irasactiveanalytics.co.uk

Website: www.irasactiveanalytics.co.uk

Testing performed at the above address only

## **DETAIL OF ACCREDITATION**

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATER		Documented In-house Methods
Normal/low level dissolved minerals (TDS <3 g/l)	Gross α activity relative to <sup>241</sup> Am (LOD to 3 Bq/kg) Gross β activity relative to <sup>40</sup> K (LOD to 5 Bq/kg)  Tritium activity (LOD to 40 MBq/kg)	LTM-GAB based on BS EN ISO 9696:2017 and BS EN ISO 9697:2019, by gas proportional counter  LTM-HTO based on BS EN ISO 9698:2019, by liquid scintillation counter
END		

Assessment Manager: RD1 Page 1 of 1