


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>UKAS MEDICAL 31263</p> <p>Accredited to ISO 15189: 2022</p>	<h3>Leading Pathology Ltd</h3> <p>Issue No: 001 Issue date: 23 March 2026</p>	
	<p>Lab F096 Building CVRM Royal Veterinary College Hawkshead Lane Hatfield AL9 7TA</p>	<p>Contact: Dr Adam Levene Tel: +44 (0) 7881057771 E-Mail: adamlevene@leadingpathology.co.uk Website: https://www.leadingpathology.co.uk/</p>
<p>Testing performed at the above address only</p>		

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>HUMAN BODY TISSUE</p> <p>Fixed and fresh tissue</p> <p>Formalin-Fixed Paraffin-Embedded (FFPE)</p> <p>In house preparations of tissue sections</p> <p>Stained slides prepared in house as above</p>	<p><u>Histological examinations for the purpose of clinical diagnosis</u></p> <p>Examination of tissues to identify or exclude morphological and cytological abnormalities for the purposes of clinical diagnosis</p> <p>Demonstration of:</p> <p>Basophilic and eosinophilic structures</p> <p>Morphological assessment/interpretation and diagnosis for Histology</p>	<p>In-house documented procedures based on equipment manuals as relevant</p> <p>Specimen dissection using LP-HT-SOP1015</p> <p>Tissue processing using Tissue Processor Diapath Donatello Series 2 LP-HT-SOP1014</p> <p>Tissue embedding using Embedding Centre Diapath Dante E.M LP-HT-SOP1006 Embedding</p> <p>Microtomy using Diapath Galileo Series 2 and Leica RM2235 SN 2936 LP-HT-SOP1005 Microtomy</p> <p>Haematoxylin & Eosin automated staining using Staining Machine-Diapath Giotto LP-HT-SOP1013 Staining Machine</p> <p>Haematoxylin & Eosin manual staining LP-HT-SOP1013</p> <p>Microscopy using Olympus microscopes</p> <p>Reporting LP-HT-SOP1012</p>
<p>END</p>		