


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 UKAS CALIBRATION 4006 Accredited to ISO/IEC 17025:2017	Guymark UK Limited Issue No: 013 Issue date: 04 December 2025	
	Geneva House International Park Starley Way Birmingham B37 7GN	Contact: Mr A Sewell Tel: +44 (0) 1384 890600 Fax: +44 (0)1384 890609 E-Mail: asew@guymark.com Website: www.guymark.com
Calibration performed by the Organisations at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Geneva House International Park Starley Way Birmingham B37 7GN Local contact Mr A Sewell	Audiometer calibrations	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Any suitable customer premises	Audiometer calibrations	Site



4006
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Guymark UK Limited
Issue No: 013 Issue date: 04 December 2025

Calibration performed by the Organisation at the locations specified

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC)	Remarks	Location Code
AUDIOMETERS: Pure tone air conduction SPL	125 Hz to 8000 Hz	0.5 dB 0.7 dB	Types 1 to 4, as classified in BS EN 60645-1:1995 (Withdrawn), BS EN 60645-1:2001 (Withdrawn), BS EN 60645-1:2015 (Withdrawn) BS EN 60645-1:2017	Lab Site
Narrowband masking	125 Hz to 8000 Hz	0.7 dB 1.0 dB		Lab Site
Pure tone bone conduction FL	250 Hz to 4000 Hz	1.50 dB		Lab & Site
Pure tone frequency	125 Hz to 8000 Hz	0.10 %		Lab & Site
Total harmonic distortion	125 Hz to 8000 Hz	0.3% THD		Lab & Site
Attenuator linearity 130 dB to 40 dB HLS <40 dB to 0 dB HLS <0 dB to -10 dB HLS	1000 Hz	0.20 dB 0.40 dB 0.50 dB	Measured between successive Hearing Level Settings	Lab
END				



4006

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Guymark UK Limited

Issue No: 013 Issue date: 04 December 2025

Calibration performed by the Organisation at the locations specified

Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of $k = 2$. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation $Q[a, b]$ stands for the root-sum-square of the terms between brackets: $Q[a, b] = [a^2 + b^2]^{1/2}$