# **Schedule of Accreditation**

issued by

**United Kingdom Accreditation Service** 

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



# Locations covered by the organisation and their relevant activities

## Laboratory locations:

| Location details   |   | Activity  | Location code |
|--|---|---|---------------|
| Address<br>19 Iris Road<br>Rogerstone<br>South Wales<br>NP10 9LE | Local contact<br>Mr Richard Bale<br>Tel: +44 (0)1633 897225<br>Fax: +44 (0)1633 897225<br>Email: info@able-calibration.co.uk<br>Website: www.able-calibration.co.uk | Electrical<br>Temperature<br>Pressure<br>Humidity | Ρ             |

# Site activities performed away from the locations listed above:

| Location details   |                  | Activity  | Location code |
|--|------------------|---|---------------|
| Customers' sites or<br>premises<br>The customers' site or<br>premises must be suitable for<br>the nature of the particular<br>calibrations undertaken and<br>will be the subject of contract<br>review arrangements<br>between the laboratory and<br>the customer. | Contact as above | Electrical<br>Temperature<br>Pressure<br>Humidity | S             |

|   | Schedule of Accreditation<br>issued by<br>United Kingdom Accreditation Service<br>2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK |
|---|--|
|   | ABLE Calibration Services Limited  |
| 4021<br>Accredited to<br>ISO/IEC 17025:2017 | Issue No: 027 Issue date: 23 June 2025   |
|   | Calibration performed by the Organisation at the locations specified   |

| Measured Quantity<br>Instrument or Gauge | Range  | Expanded<br>Measurement<br>Uncertainty (k = 2)   | Remarks   | Location<br>Code |
|--|--|--|---|------------------|
| ELECTRICAL                               |  |  | Electrical calibrations<br>are performed as a<br>comparison against a<br>reference standard |                  |
| DC VOLTAGE                               | 0 mV to 100 mV<br>0.1 V to 1000 V  | 5.0 μV<br>40 μV/V  | These values can be generated or measured   | P&S              |
| DC HIGH VOLTAGE<br>Measurement only      | 1 kV to 3 kV<br>3 kV to 30 kV  | 1.7 % + 20 V<br>0.43 % + 20 V  | For the calibration of<br>high voltage sources  |                  |
| DC CURRENT                               | 1.0 μA to 100 μA<br>100 μA to 1 mA   | 500 μA/A + 10 nA<br>500 μA/A   | These values can be generated or measured   | P&S              |
|  | 1 mA to 10 mA<br>10 mA to 100 mA<br>100 mA to 1 A<br>1 A to 3 A<br>3 A to 10 A<br>10 A to 20 A   | 200 μΑ/Α<br>200 μΑ/Α<br>300 μΑ/Α<br>300 μΑ/Α<br>0.050 %<br>0.25 % + 6 mA   |   |                  |
| DC CURRENT<br>Simulation                 | 10 A to 1000 A   | 0.40 % + 6 mA  | 50 turn coil – for the<br>calibration of<br>clampmeters                                     |                  |
| DC RESISTANCE                            | $\begin{array}{l} 0 \ m\Omega \\ 1 \ m\Omega \\ 10 \ m\Omega \\ 100 \ m\Omega \\ 100 \ m\Omega \\ \tauo \ 100 \ \Omega \\ 100 \ \Omega \ to \ 1 \ M\Omega \\ 100 \ \Omega \ to \ 1 \ M\Omega \\ 1 \ M\Omega \ to \ 100 \ M\Omega \\ 100 \ M\Omega \ to \ 1000 \ M\Omega \end{array}$ | 0.03 m $\Omega$<br>0.03 m $\Omega$<br>0.02 m $\Omega$<br>0.2 m $\Omega$<br>50 $\mu\Omega/\Omega$ + 1 m $\Omega$<br>50 $\mu\Omega/\Omega$<br>130 $\mu\Omega/\Omega$<br>0.40 %<br>0.50 % | These values can be<br>generated or measured,<br>except where stated                        | P&S              |
| (Specific Value)<br>Generation only      | 100 Ω  | 24 μΩ/Ω  |   |                  |
| AC RESISTANCE                            | 1 kHz<br>1 Ω to 10 kΩ  | 0.10 %   |   | P & S            |
| AC VOLTAGE                               | 1 mV to 100 mV<br><i>40 Hz to 1 kHz</i>  | 500 μV/V   | These values can be generated or measured   | P & S            |
|  | 0.1 V to 1 V<br>40 Hz to 1 kHz<br>1 kHz to 50 kHz  | 400 μV/V<br>600 μV/V   |   |                  |
|  |  |  |   |                  |

# Calibration and Measurement Capability (CMC)

|  | Schedule of Accreditation<br>issued by<br>United Kingdom Accreditation Service<br>2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK |  |  |
|--|--|--|--|
| 4021<br>Accredited to<br>ISO/IEC 17025:2017                          | ABLE Calibration Services Limited<br>Issue No: 027 Issue date: 23 June 2025  |  |  |
| Calibration performed by the Organisation at the locations specified |  |  |  |

| Measured Quantity<br>Instrument or Gauge | Range   | Expanded<br>Measurement<br>Uncertainty (k = 2)           | Remarks   | Location<br>Code |
|--|---|--|---|------------------|
| ELECTRICAL (continued)                   |   |  |   |                  |
| AC VOLTAGE (continued)                   | 1 V to 10 V<br>40 Hz to 1 kHz<br>1 kHz to 50 kHz  | 400 μV/V<br>600 μV/V                                     |   | P&S              |
|  | 10 V to 100 V<br>40 Hz to 1 kHz<br>1 kHz to 50 kHz  | 400 μV/V<br>600 μV/V                                     |   |                  |
|  | 100 V to 750 V<br>40 Hz to 1 kHz  | 500 μV/V   |   |                  |
| AC HIGH VOLTAGE<br>Measurement only      | 50 Hz<br>750 V to 10 kV   | 1.2 % + 20 V   | For the calibration of<br>high voltage sources  | P & S            |
| AC CURRENT                               | 40 Hz to 1 kHz<br>1 μA to 100 μA<br>100 μA to 1 mA<br>1 mA to 10 mA<br>10 mA to 100 mA<br>100 mA to 1 A<br>1 A to 3 A | 0.16 %<br>0.11 %<br>0.11 %<br>0.11 %<br>0.10 %<br>0.14 % | These values can be generated or measured   | P&S              |
|  | 40 Hz to 100 Hz<br>3 A to 10 A<br>10 A to 20 A  | 0.10 %<br>0.43 % + 10 mA                                 |   |                  |
| Simulation                               | <i>40 to 50 Hz</i><br>10 A to 1000 A  | 0.50 % + 10 mA   | 50 turn coil – for the<br>calibration of<br>clampmeters   | P&S              |
| FREQUENCY                                |   |  |   | P & S            |
| Measurement                              | 5 Hz to 25 MHz<br>25 MHz to 1.3 GHz   | 2.8 μHz/Hz + 16 μHz<br>2.8 μHz/Hz + 160 mHz              | Generation up to<br>15 MHz<br>May be expressed as<br>events per unit time for<br>devices such as<br>Tachometers and 1/f for<br>oscilloscope period<br>calibration |                  |
|  |   |  |   |                  |
|  |   |  |   |                  |
|  |   |  |   |                  |

| 4021<br>Accredited to<br>ISO/IEC 17025:2017 | Schedule of Accreditation<br>issued by<br>United Kingdom Accreditation Service<br>2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK<br>ABLE Calibration Services Limited<br>Issue No: 027 Issue date: 23 June 2025 |  |   |                  |  |
|---|---|--|---|------------------|--|
|   | Calibration performed by the O  | rganisation at the locations s                 | pecified  |                  |  |
| Measured Quantity<br>Instrument or Gauge    | Range   | Expanded<br>Measurement<br>Uncertainty (k = 2) | Remarks   | Location<br>Code |  |
| TIME INTERVAL                               | 1.0 s to 6 Hours  | 70 ms  | For the calibration of stopwatches and timing devices | P & S            |  |
| Revolutions Per Minute                      | 3 RPM to 90 000 RPM   | 1.0 RPM  | For the calibration of<br>Optical Tachometers         |                  |  |
| CAPACITANCE                                 | <i>1 kHz</i><br>100 pF to 100 μF  | 0.10 %   |   |                  |  |
| INDUCTANCE                                  | <i>1 kHz</i><br><i>10.</i> μF to 1 H<br>1 H to 10 H   | 0.10 %<br>0.30 %                               |   |                  |  |
| SIMULATION                                  |   |  |   | P & S            |  |
| Thermocouple simulators and                 |   |  |   |                  |  |

| indicators, calibration by electrical simulation                |  |                                 |  |     |
|---|--|---------------------------------|--|-----|
| Noble metal thermocouples<br>Type R & S                         | -50 °C to +1768 °C   | 1.0 °C                          | With cold junction   |     |
|   | -50 °C to +200 °C<br>200 °C to 1768 °C                         | 1.0 °C<br>0.50 °C               | Without cold junction compensation   |     |
| Base metal thermocouples<br>Type K, J, T & N                    | -200 °C to +1372 °C  | 0.70 °C                         | With cold junction   |     |
|   | -200 °C to 0 °C<br>0 °C 1372 °C                                | 0.30 °C<br>0.15 °C              | compensation<br>Without cold junction<br>compensation  |     |
| Cold junction compensation                                      | At ambient temperature   | 0.32 °C                         |  |     |
| Resistance thermometers (PT100)                                 | -200 °C to +800 °C   | 0.10 °C                         |  |     |
| PRESSURE  |  |                                 |  |     |
| Gas pressure (gauge)  |  |                                 | Methods consistent with EURAMET CG3.   |     |
| Calibration of pressure<br>indicating instruments and<br>gauges | -90 kPa to 200 kPa<br>200 kPa to 700 kPa<br>700 kPa to 3.5 MPa | 0.24 kPa<br>0.60 kPa<br>2.3 kPa | Instruments with an<br>electrical output can be<br>calibrated.<br>Absolute pressures can<br>be generated across<br>these gauge ranges<br>which will attract an<br>additional uncertainty of<br>0.33 kPa. | P&S |
| 1   |  |                                 |  |     |

| 4021<br>Accredited to<br>ISO/IEC 17025:2017   | Schedule of Accreditation<br>issued by<br>United Kingdom Accreditation Service<br>2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK<br>ABLE Calibration Services Limited<br>Issue No: 027 Issue date: 23 June 2025 |  |   |                  |
|---|---|--|---|------------------|
|   | Calibration performed by the Org  | janisation at the locations s  | pecified  |                  |
| Measured Quantity<br>Instrument or Gauge  | Range   | Expanded<br>Measurement<br>Uncertainty (k = 2)   | Remarks   | Location<br>Code |
| Gas pressure (absolute)<br>Calibration of pressure<br>indicating instruments and<br>gauges  | 10 kPa to 200 kPa   | 0.33 kPa   | Instruments with an<br>electrical output can be<br>calibrated.  | P&S              |
| PRESSURE (continued)<br>Hydraulic pressure (gauge)<br>Calibration of pressure<br>indicating instruments and<br>gauges   | 0 MPa to 40 MPa   | 46 kPa   | Instruments with an<br>electrical output can be<br>calibrated.<br>Absolute pressures can<br>be generated across<br>this gauge range which<br>will attract an additional<br>uncertainty of 0.33 kPa. | P&S              |
| TEMPERATURE<br>Temperature indicators and<br>recorders, with temperature<br>sensor(s)   | -90 °C to -70 °C<br>-70 °C to -50 °C<br>-50 °C to +140 °C<br>140 °C to 200 °C<br>200 °C to 400 °C<br>400 °C to 600 °C<br>-90 °C to -50 °C<br>-50 °C to +140 °C<br>140 °C to 200 °C<br>200 °C to 400 °C                          | 0.31 °C<br>0.15 °C<br>0.080 °C<br>0.15 °C<br>0.80 °C<br>0.90 °C<br>0.32 °C<br>0.13 °C<br>0.15 °C<br>0.15 °C<br>0.15 °C | Instruments with an<br>electrical output can be<br>calibrated.<br>Calibrations performed<br>within Dry media<br>Calibrations performed<br>within dry media  | P&S              |
| Calibration in ambient air<br>Temperature controlled<br>chambers, fridges, freezers,<br>ovens, furnaces and liquid<br>baths (inclusive of associated<br>indicators, controllers and<br>recorders, all with sensors,<br>within the specified<br>parameters and ranges) | 200 °C to 400 °C<br>400 °C to 600 °C<br>10 °C to 30 °C<br>-90 °C to -70 °C<br>-70 °C to -50 °C<br>-50 °C to +200 °C<br>200 °C to 400 °C<br>400 °C to 600 °C   | 0.80 °C<br>0.90 °C<br>0.40 °C<br>0.30 °C<br>0.15 °C<br>0.050 °C<br>0.80 °C<br>1.5 °C                                   | Single or dual<br>monitoring probes.<br>Time dependent<br>temperature profiling,<br>also referred to as<br>spatial temperature<br>surveying or mapping  | S                |

1.3 °C 1.2 °C 1.4 °C 1.8 °C

-80 °C to -50 °C -50 °C to +200 °C 200 °C to 400 °C 400 °C to 600 °C Multipoint monitoring probes. Time dependent temperature profiling, also referred to as spatial temperature surveying or mapping

|  | Schedule of Accreditation<br>issued by<br>United Kingdom Accreditation Service<br>2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK |   |                            |          |  |
|--|--|---|----------------------------|----------|--|
| 4021<br>Accredited to<br>ISO/IEC 17025:2017                          | ABL  | E Calibration Servic<br>No: 027 Issue date: | es Limited<br>23 June 2025 |          |  |
| Calibration performed by the Organisation at the locations specified |  |   |                            |          |  |
| Measured Quantity  | Range  | Expanded                                    | Remarks                    | Location |  |

| Measured Quantity<br>Instrument or Gauge             | Range   | Expanded<br>Measurement<br>Uncertainty (k = 2)       | Remarks  | Location<br>Code |
|--|---|--|--|------------------|
| Temperature block calibrators                        | -90 °C to -70 °C<br>-70 °C to -50 °C<br>-50 °C to +200 °C<br>200 °C to 400 °C<br>400 °C to 600 °C | 0.30 °C<br>0.15 °C<br>0.040 °C<br>0.50 °C<br>0.60 °C | Method consistant with<br>Euramet CG13   | P&S              |
| HUMIDITY   |   |  |  | P & S            |
| Relative humidity instruments in ambient air         | 11 %rh to 50 %rh<br>50 %rh to 95 %rh<br><i>at 15 ℃ to 30 ℃</i>                                    | 1.5 %rh<br>2.0 %rh                                   | Single point calibration<br>in prevailing conditions<br>using reference<br>instrument. |                  |
| Relative humidity instruments at ambient temperature | 11 %rh<br>35 %rh<br>50 %rh<br>80 %rh<br>at 15 ℃ to 30 ℃   | 0.39 %rh<br>0.68 %rh<br>1.0 %rh<br>1.4 %rh           | Calibration using reference salts.   |                  |
| Temperature probes built into humidity instruments   | 15 °C to 30 °C  | 0.40 °C  | Comparison with reference device.  |                  |
| END  |   |  |  |                  |



### Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where *q* is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$