Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

4042

Accredited to ISO/IEC 17025:2017

Normec Westfield Limited

Issue No: 021 **Issue date:** 19 November 2026

4 Mollins Court
Cumbernauld
G68 9HP

Contact: Mr Jim Bowman
Tel: +44 (0)1236 786300
E-Mail: jimbowman@normecgroup.com
Website: www.normecgroup.com

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS	Microbiology Testing:	In-House Documented Methods based on the Microbiology of Drinking Water (MDW)
	Enumeration of:	
WATERS – drinking (non regulatory), recreational, surface and process	Total Aerobic Colony Count at 22°C and 37°C	1) LO 001 based on MDW, Part 7, (2020) 2) LO 002 (extended dilution method) based on MDW, Part 7, (2020)
Endoscope Rinse waters	Total Aerobic Colony Count	LO 040, in-house method using membrane filtration onto TSA based on HTM 2030 (withdrawn), section 9.223
WATERS – drinking (non regulatory), recreational, surface and process	Coliform bacteria	LO 004 based on MDW, Part 4B (2016)
	Escherichia coli	LO 004 based on MDW, Part 4B (2016)
	Pseudomonas aeruginosa	LO 005 based on MDW, Part 8, (2015)
	Enterococci	LO 028 based on MDW, Part 5, (2012)
	Sulphite Reducing Clostridia	LO 038 based on MDW, Part 6, (2021)

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS (cont'd)	Microbiology Testing: (cont'd) Enumeration of: (cont'd)	In-House Documented Methods
WATERS – drinking (non regulatory), recreational, surface and process	*Legionella pneumophila* 1, 2-14 and *Legionella* species	LO 009 based on ISO 11731:2017 using filtration with washing, using GVPC, identification using latex agglutination [Matrix A; Procedure 8, 9 & 10, Media C] For turbid samples and samples with high deposit where concentration by filtration is not possible; LO 009 based on ISO 11731:2017 with concentration by centrifugation, identification using latex agglutination [Matrix A, B & C]
WATERS – drinking (non regulatory) and process (closed system)	*Pseudomonas* species (presumptive) Total Aerobic Colony Count at 30ºC	LO 045 by membrane filtration technique using CFC agar, confirmed by oxidase reaction 1) LO 001 based on MDW, Part 7, 2020 2) LO 002 (extended dilution method) based on MDW, Part 7, 2020
ENVIRONMENTAL SAMPLES	Incubation and Enumeration of: Contact plates Settle plates Active air sampling plates	In-House Method LO 003 for the recovery of: 1) Total aerobic bacteria count (TSA medium incubated at 37°C for 2 days) 2) Total moulds/yeast count (SDA medium incubated at 22°C for 3 days)
END		