

**issued by**

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



**Accredited to  
ISO/IEC 17025:2017**

**Issue No: 006    Issue date: 25 January 2025**

**The Electrical Measurements  
Laboratory  
St Leonards Building  
Redhill Business Park  
Stafford  
Staffordshire  
ST16 1WT**

**Contact: Mr N J Podmore**  
**Tel: +44 (0)1785 786397**  
**E-Mail: [neil.podmore@ge.com](mailto:neil.podmore@ge.com)**  
**Website: [www.gegridsolution.com](http://www.gegridsolution.com)**

**Calibration performed at the above address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
Values and uncertainties listed below are applicable for the calibration of both measuring instruments and for instruments with an output. The method used is by direct comparison unless otherwise stated in the remarks column.			
<b>ELECTRICAL</b>			
DC Resistance	Specific values as shunts		<p>The value of a 4-terminal shunt may be expressed as a terminal voltage at a specific current</p> <p>Supplies up to 1 kA are available to the laboratory</p> <p>Test current 70 A Test current 14 A</p> <p>Resistors of modest dimensions, suitable for oil immersion, can be measured over the temperature range 15 °C to 30 °C</p>
	1 mΩ 10 mΩ 100 mΩ  at negligible power  1 mΩ 10 mΩ 100 mΩ  1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ  100 kΩ 1 MΩ 10 MΩ 100 MΩ	56 μΩ/Ω 26 μΩ/Ω 24 μΩ/Ω   30 μΩ/Ω 23 μΩ/Ω 10 μΩ/Ω  7.0 μΩ/Ω 3.0 μΩ/Ω 2.0 μΩ/Ω 2.0 μΩ/Ω 4.0 μΩ/Ω  8.0 μΩ/Ω 14 μΩ/Ω 19 μΩ/Ω 38 μΩ/Ω	
Other values	10 Ω to 100 MΩ	200 μΩ/Ω	
<b>ELECTRICAL</b> (continued)			



**4085**  
Accredited to  
**ISO/IEC 17025:2017**

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**UK Grid Solutions Limited**  
**Issue No: 006 Issue date: 25 January 2025**

**Calibration performed at main address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
DC Voltage	1 V, 10 V  0 V to 10 V 10 V to 200 V 200 V to 1500 V  1.5 kV to 5 kV	0.80 $\mu\text{V/V}$  1.0 $\mu\text{V/V} + 0.60 \mu\text{V}$ 5.0 $\mu\text{V/V}$ 8.0 $\mu\text{V/V}$  0.050 % + 0.60 V	
DC Current	0 A to 200 $\mu\text{A}$ 200 $\mu\text{A}$ to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 1 A  1 A to 14 A 14 A to 100 A	8.0 $\mu\text{A/A} + 0.50 \text{ nA}$ 11 $\mu\text{A/A} + 2.5 \text{ nA}$ 11 $\mu\text{A/A} + 25 \text{ nA}$ 12 $\mu\text{A/A} + 0.25 \mu\text{A}$ 25 $\mu\text{A/A} + 2.5 \mu\text{A}$  20 $\mu\text{A/A}$ 41 $\mu\text{A/A}$	Supplies available up to 100 A
DC Power	0 W to 150 kW	The power uncertainty will be the sum of the individual uncertainties for the corresponding voltage and current measurements as they are correlated terms	Limiting voltage 1.5 kV Limiting current 100 A
AC Voltage	10 mV to 210 mV 40 Hz to 20 kHz 20 kHz to 100 kHz  210 mV to 2.1 V 40 Hz to 20 kHz 20 kHz to 100 kHz  2.1 V to 21 V 40 Hz to 20 kHz 20 kHz to 100 kHz  21 V to 210 V 40 Hz to 20 kHz 20 kHz to 100 kHz  210 V to 300 V 10 kHz to 100 kHz  210 V to 1010 V 40 Hz to 10 kHz  50 Hz to 60 Hz 1010 V to 10 kV 1 kV to 5 kV	300 $\mu\text{V/V} + 12 \mu\text{V}$ 500 $\mu\text{V/V} + 12 \mu\text{V}$  51 $\mu\text{V/V} + 12 \mu\text{V}$ 150 $\mu\text{V/V} + 12 \mu\text{V}$  32 $\mu\text{V/V} + 120 \mu\text{V}$ 56 $\mu\text{V/V} + 120 \mu\text{V}$  42 $\mu\text{V/V} + 1.2 \text{ mV}$ 72 $\mu\text{V/V} + 1.2 \text{ mV}$  95 $\mu\text{V/V} + 12 \text{ mV}$  45 $\mu\text{V/V} + 12 \text{ mV}$  0.90 % + 5.8 V 0.16 % + 0.30 V	
<b>ELECTRICAL</b> (continued)			



4085  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**UK Grid Solutions Limited**  
**Issue No: 006 Issue date: 25 January 2025**

**Calibration performed at main address only**

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
AC Current	<i>40 Hz to 1 kHz</i> 30 $\mu$ A to 210 $\mu$ A 210 $\mu$ A to 2.1 mA 2.1 mA to 21 mA 21 mA to 210 mA 210 mA to 2.1 A  <i>45 Hz to 60 Hz</i> 2.5 A to 1000 A	200 $\mu$ A/A 60 $\mu$ A/A + 100 nA 63 $\mu$ A/A + 240 nA 75 $\mu$ A/A + 2.4 $\mu$ A 120 $\mu$ A/A + 26 $\mu$ A  0.030 %	Supplies available up to 500 A
AC Current Ratio			
Primary Currents	<i>50 Hz</i> 0.25, 0.5, 1, 2.5, 5, 10, 25, 50, 100, 250, 500, and 1000 A  Secondary Current 5A only	0.0150 % ratio 0.75 minutes of phase (for CTs of accuracy class 0.5 or better)  0.080 % ratio 2.5 minutes of phase	For calibration of current transformers with resistive burdens up to 10 VA  Supplies available up to 500 A For CTs of accuracy classes 0.5 to 5)
AC Power			
at unity power factor	<i>50 Hz to 60 Hz</i> Maximum voltage 1 kV 2.5 W to 6 kW 6 kW to 100 kW	0.0080 % 0.014 %	Maximum current 6 A Maximum current 100 A  The maximum current and voltage are for a single-phase system. For a three-phase system they will be reduced to 25 A and 500 V respectively.
Phase Angle	0° to 360° <i>50 Hz to 5 kHz</i> <i>5 kHz to 50 kHz</i> <i>50 kHz to 100 kHz</i>	0.020° 0.040° 0.060°	Equal amplitudes in the range 50 mV to 120 V. Current range 0.5 A to 5 A at 50 to 60 Hz. Increased uncertainties will apply for unequal amplitudes. The uncertainties relate to the calibration of a phase meter of suitable resolution and stability.
Frequency			
Specific values	1 Hz, 10 Hz, 50 Hz, 100 Hz, 1 kHz, 10 kHz, 100 kHz, 1 MHz, 5 MHz and 10 MHz	8.1 in $10^{12}$	
Time Interval	100 $\mu$ s to 10 minutes	8.0 in $10^{12}$ + 1.0 ns	



4085  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**UK Grid Solutions Limited**  
**Issue No: 006   Issue date: 25 January 2025**

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
Capacitance	100 pF to 1 nF 100 Hz 1 kHz to 10 kHz	0.22 % 0.080 %	Fixed capacitors are available within this range for the calibration of bridges and capacitance meters.
	1 nF to 10 nF 50/60 Hz 100 Hz to 10 kHz	0.28 % 0.031 %	
	10 nF to 100 nF 50/60 Hz 100 Hz to 10 kHz	0.066 % 0.034 %	
	100 nF to 1 $\mu$ F 50 Hz to 10 kHz	0.058 %	
	1 $\mu$ F to 10 $\mu$ F 50 Hz to 1 kHz	0.066 %	
Temperature indicators, calibration by electrical simulation			Including unit under test Reference junction compensation
Thermocouple Indicators			
Type K	-200 °C to 0 °C	0.40 °C	
	0 °C to 1300 °C	0.44 °C	
Type T	-200 °C to 0 °C	0.40 °C	
	0 °C to 400 °C	0.44 °C	
Reference junction measurement	Ambient range 18 °C to 22 °C	0.10 °C	
END			



4085  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**UK Grid Solutions Limited**  
**Issue No: 006   Issue date: 25 January 2025**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$