


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>4211</b></p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>NHS Greater Glasgow &amp; Clyde, Microbiology Department</b></p> <p>Issue No: 013 Issue date: 10 February 2025</p>	
	<p>Clinical Microbiology Department Level 4 New Lister Building Glasgow Royal Infirmary 10-16 Alexandra Parade Glasgow G31 2ER</p>	<p>Contact: Ms Stephanie McPhee Tel: +44 (0)141 2018570 Fax: +44 (0)141 2018535 E-Mail: Stephanie.McPhee@ggc.scot.nhs.uk</p>
<p>Testing performed at the above address only</p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>WATER (Drinking, ground, surface and process waters)</p> <p>Drinking and recreation waters (man made)</p> <p>Healthcare waters (renal dialysis water)</p>	<p><u>Microbiological Tests</u></p> <p>Enumeration of:</p> <p><i>Legionella</i> spp and subsequent identification of: <i>Legionella pneumophila</i> serogroup 1 <i>L. pneumophila</i> serogroups 2-14 <i>Legionella</i> species</p>	<p>Documented In-House method LP532 based on ISO 11731:2017 for Matrix A and B, following procedures 8, 9 and 10 using Medium C</p>
	<p>Total Viable Count</p>	<p>Documented In-House method LP538 based on the Microbiology of Drinking Water Part 7 2020</p>
	<p>Coliforms &amp; <i>Escherichia coli</i></p>	<p>Documented In-House method LP538 based on ISO 9308-01:2014 (E)</p>
	<p><i>Pseudomonas</i> spp and <i>Pseudomonas aeruginosa</i></p>	<p>Documented In-House method LP538 based on the Microbiology of Drinking Water Part 8 2015</p>
	<p>Total Viable Count</p>	<p>Documented In-House method LP538 based on the Microbiology of Drinking Water Part 7 2020, the Renal Association Guidelines and ISO 23500:2024</p>
	<p>Coliforms &amp; <i>Escherichia coli</i></p>	<p>Documented In-House method LP538 based on ISO 9308-01:2014 (E), the Renal Association Guidelines and ISO 23500:2024</p>



4211  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Microbiology Department Glasgow Royal Infirmary**

**Issue No: 013 Issue date: 10 February 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
WATERS	<u>Microbiological Tests</u> (cont'd)	
Healthcare waters (renal dialysis water) (cont'd)	Enumeration of:	Documented In-House method LP538 based on the Microbiology of Drinking Water Part 8 2015, the Renal Association Guidelines and ISO 23500:2024
Cardiothoracic Heater-cooler water	Total Viable Count	Documented In-House method LP538 based on the Microbiology of Drinking Water Part 7 2020 and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019
	Coliforms & <i>Escherichia coli</i>	Documented In-House method LP538 based on ISO 9308-01:2014 (E) and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019
	<i>Pseudomonas</i> spp and <i>Pseudomonas aeruginosa</i>	Documented In-House method LP538 based on the Microbiology of Drinking Water Part 8 2015 and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019
Endoscopy rinse waters	Total Viable Count	Documented In-House method LP538 based on the the requirements of HTM -01-06
	<i>Pseudomonas aeruginosa</i>	Documented In-House method LP538 based on the the requirements of HTM -01-06
	Atypical Mycobacteria	Documented In-House method LP538 based on the requirements of HTM -01-06



4211  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Microbiology Department Glasgow Royal Infirmary**

**Issue No: 013 Issue date: 10 February 2025**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
ENVIRONMENTAL SAMPLES	<u>Microbiological Tests</u> (cont'd)  Incubation and Enumeration of: Total Viable Count Yeasts and Moulds  Sampling for Contact Plates, Settle Plates, Active Air Sampling Plates  <u>Physical Tests</u>  Airborne Particle Concentration	Settle plates, contact plates, active air-sampled plates following documented In-House method LP539  Documented In-House method LP539 at specified customer locations using Quaria Micro-Flow active air sampler  Documented In-House method LP539 at specified customer locations using Hach Met One 3400 particle counter
END		