


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 4211 Accredited to ISO/IEC 17025:2017	NHS Greater Glasgow & Clyde, Microbiology Department	
	Issue No: 014 Issue date: 18 September 2025	
	Environmental Microbiology Department Level 4 New Lister Building Glasgow Royal Infirmary 10-16 Alexandra Parade Glasgow G31 2ER	Contact: Stephanie McPhee Tel: +44 (0)141 2018570 Fax: +44 (0)141 2018535 E-Mail: Stephanie.McPhee@nhs.scot
Testing performed by the Organisation at the locations specified		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code	
Address Level 4 New Lister Building Glasgow Royal Infirmary 10-16 Alexandra Parade Glasgow G31 2ER	Local contact Stephanie McPhee	Testing: Microbiological testing	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Hospital wards, theatres and laboratory facilities	Sampling of air and surfaces at clients' premises Testing air quality for physical quality at clients' premises	Site



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Greater Glasgow & Clyde, Microbiology Department

Issue No: 014

Issue date: 18 September 2025

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATER (Drinking, ground, surface and process waters)	<u>Microbiological Tests</u> Enumeration of: <i>Legionella</i> spp and subsequent identification of: <i>Legionella pneumophila</i> serogroup 1 and 2-14 <i>Legionella</i> spp	Documented In-House methods LP532 based on ISO 11731:2017 for Matrix A and B, following procedures 8, 9 and 10 using Medium C, confirmed using latex agglutination	A
	Drinking and recreation waters (man-made)	Total Viable Count Coliforms & <i>Escherichia coli</i> <i>Pseudomonas</i> spp and <i>Pseudomonas aeruginosa</i>	A A A
Healthcare waters (renal dialysis water)	Total Viable Count	LP538 based on the Microbiology of Drinking Water Part 7 2020, the Renal Association Guidelines and ISO 23500:2024	A
	Coliforms & <i>Escherichia coli</i>	LP538 based on ISO 9308-01:2014 (E), the Renal Association Guidelines and ISO 23500:2024	A
	<i>Pseudomonas</i> spp and <i>Pseudomonas aeruginosa</i>	LP538 based on the Microbiology of Drinking Water Part 8 2015, the Renal Association Guidelines and ISO 23500:2024	A



4211

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Greater Glasgow & Clyde, Microbiology Department

Issue No: 014

Issue date: 18 September 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATER (Healthcare cardiothoracic heater-cooler water)	<u>Microbiological Tests</u> (cont'd)	Documented In-House methods	
	Enumeration of: Total Viable Count	LP538 based on the Microbiology of Drinking Water Part 7 2020 and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019	A
	Coliforms & <i>Escherichia coli</i>	LP538 based on ISO 9308-01:2014 (E) and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019	A
Endoscopy rinse waters	<i>Pseudomonas</i> spp and <i>Pseudomonas aeruginosa</i>	LP538 based on the Microbiology of Drinking Water Part 8 2015 and NHS Scotland Guidance for Decontamination and testing of Cardiac Heater-Cooler Units V1 2019	A
	Total Viable Count	LP538 based on the requirements of HTM -01-06	A
	<i>Pseudomonas aeruginosa</i>	LP538 based on the requirements of HTM -01-06	A
	Atypical Mycobacteria	LP538 based on the requirements of HTM -01-06	A



4211

Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

NHS Greater Glasgow & Clyde, Microbiology Department

Issue No: 014 Issue date: 18 September 2025

Testing performed by the Organisation at the locations specified

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
ENVIRONMENTAL SAMPLES	<u>Microbiological Tests</u> (cont'd)	Documented In-House methods	
	Incubation and Enumeration of: Contact Plates (including finger dabs) Settle Plates Active Air Sampling Plates	LP539 for TSA and SDA contact and active air-sampled plates incubated as agreed with customer	A
	Sampling for Contact Plates, Settle Plates and Active Air Sampling Plates	LP539 at specified customer locations using TSA and SDA for surface contact plates and air-sampled plates using Quaria Micro-Flow active air sampler	Site
	<u>Physical Tests</u>		
	Airborne Particle Concentration	LP539 at specified customer locations using Hach Met One 3400 particle counter	Site
END			