


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 4346 Accredited to ISO/IEC 17025:2017	RWE Generation UK plc	
	Issue No: 009 Issue date: 05 September 2022	
	Didcot B Power Station Didcot Oxfordshire OX11 7YU	This laboratory does not accept commercial or external enquiries Website: www.rwe.com
Testing performed by the Organisation at the locations specified below		

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address Didcot B Power Station Didcot Oxfordshire OX11 7YU Local contact This laboratory does not accept commercial or external enquiries Website: www.rwe.com	Support Functions: Quality System Quality Audit Administration Sampling and Testing: Chemical Tests	A

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Consented water discharges for RWE Generation PLC, Didcot	Environmental Sampling	B



4346
Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

RWE Generation UK plc
Issue No: 009 Issue date: 05 September 2022

Testing performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used	Location Code
WATERS - River water and site industrial effluent	<u>Sampling for chemical analysis conducted in-house</u>	Documented In-House Methods	B
	<u>Chemical Analysis</u>		
	Total Chlorine (0 - 5 mg/l)	Documented In-House Method to BS EN ISO 7393-2:2000 (DID-LAB-008)	A
	Oil in Water (0 - 200 mg/l)	Documented In-House Method using Partition Infrared based on standard method 5520C for examination of water and waste water (DID-LAB-009)	A
	pH (0 - 14)	Documented In-House Method to BS 6068-2.5:1995 (DID-LAB-006)	A
	Suspended solids (≥ 2 mg/l)	Documented In-House Method to BS EN 872:2005 (DID-LAB-010)	A
	Chloride Sulphate	Documented In-House Method to BS EN ISO 10304-1:1995 (DID-LAB-007)	A
END			