

Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Accredited to
ISO/IEC 17025:2017

Randox Laboratories Limited, Trading as Randox Testing Services

Issue No: 041 Issue date: 23 January 2026

34 Diamond Road Crumlin Co Antrim BT29 4QY	Contact: Ms Emma McGoldrick Tel: +44 (0)2894 422413 Fax: +44 (0)1617 412777 E-Mail: emma.mcgoldrick@randox.com Website: www.randoxtestingservices.com
---	--

Testing performed at the above address only

DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS Urine (Preserved and unpreserved)	<u>Workplace Drug Analysis and Medical Legal Analysis</u> Presumptive Screening for the presence of drugs of abuse (DOAI and DOA II) (cut-off limit): Amphetamines (500 ng/mL) Methamphetamine (500 ng/mL) Benzodiazepines (200 ng/mL) Barbiturates (200 ng/mL) Opiates (300 ng/mL) Cocaine metabolites (150 ng/mL) Methadone (300 ng/mL) Cannabinoids (50 ng/mL) Phencyclidine (25 ng/mL) Ketamine (75ng/ml) Oxycodone Class 1 and Class 2 (100ng/ml) Generic opioids (100ng/ml) Fentanyl (1ng/ml) Propoxyphene (300ng/ml) Methaqualone (300ng/ml)	Documented in house method SOP RTL 2213 RTL using Randox Evidence Investigator (method reference EI) biochip assay



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,
Trading as Randox Testing Services**

Issue No: 041 Issue date: 23 January 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS (cont'd)	<u>Workplace Drug Analysis and Medical Legal Analysis (cont'd)</u>	
Urine (Preserved and unpreserved)	Presumptive Screening for the presence of drugs of abuse (cont'd) (DOAI and DOA II) (cut-off limit): Buprenorphine (2ng/ml) LSD (0.5ng/ml) MDMA Group (500ng/ml)	Documented in house method SOP RTL 2213 RTL using Randox Evidence Investigator (method reference EI) biochip assay
Urine (Preserved and unpreserved)	Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]: Δ9-Tetrahydrocannabinoic acid (THC-COOH) (15ng/ml) [5-150ng/ml]	Documented in house method A4URN and WI0227-RTL using solid phase extraction LC-MS/MS
Urine (Preserved and Unreserved)	Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]: Amphetamines:- Amphetamine (200ng/ml) [30-800ng/ml] Methamphetamine (200ng/ml) [30-800ng/ml] MDMA (200ng/ml) [30-800ng/ml] MDA (200ng/ml) [30-800ng/ml]	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,
Trading as Randox Testing Services**

Issue No: 041 Issue date: 23 January 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS (cont'd) Urine (Preserved and Unreserved)	<p><u>Workplace Drug Analysis and Medical Legal Analysis (cont'd)</u></p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Cocainics:- Benzoyllecgonine (100ng/ml) [30-800ng/ml] Cocaine (100ng/ml) [30-800ng/ml]</p> <p>Opiates:- 6-monoacetylmorphine (10ng/ml) [3-80ng/ml] Morphine (300ng/ml) [30-800ng/ml] Codeine (300ng/ml) [30-800ng/ml] Dihydrocodeine (300ng/ml) [30-800ng/ml]</p> <p>Opioids:- Methadone (250ng/ml) [30-800ng/ml] EDDP (75ng/ml) [30-800ng/ml] Oxycodone (100ng/ml) [30-800ng/ml] Oxymorphone (100ng/ml) [30-800ng/ml]</p>	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,
Trading as Randox Testing Services**

Issue No: 041 Issue date: 23 January 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS (cont'd) Urine (Preserved and Unreserved)	<p><u>Workplace Drug Analysis and Medical Legal Analysis (cont'd)</u></p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Opioids (cont'd):- Hydrocodone (100ng/ml) [30-800ng/ml] Hydromorphone (100ng/ml) [30-800ng/ml] Fentanyl (1ng/ml) [0.3-8.0ng/ml] Tramadol (100ng/ml) [30-800ng/ml]</p> <p>z-Drugs:- Zolpidem (10ng/ml) [3-80ng/ml]</p> <p>Ketamine:- Ketamine (75ng/ml) [30-800ng/ml] Norketamine (75ng/ml) [30-800ng/ml]</p>	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,
Trading as Randox Testing Services**

Issue No: 041 Issue date: 23 January 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS (cont'd) Urine (Preserved and Unreserved)	<u>Workplace Drug Analysis and Medical Legal Analysis (cont'd)</u> Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]: Benzodiazepines:- Alpha-hydroxyalprazolam (100ng/ml) [30-800ng/ml] 7-aminoclonazepam (100ng/ml) [30-800ng/ml] 7-aminoflunitrazepam (100ng/ml) [30-800ng/ml] 7-aminonitrazepam (100ng/ml) [30-800ng/ml] Bromazepam (100ng/ml) [30-800ng/ml] Desalkylflurazepam (100ng/ml) [30-800ng/ml] Diazepam (100ng/ml) [30-800ng/ml] Lorazepam (100ng/ml) [30-800ng/ml] Lormetazepam (100ng/ml) [30-800ng/ml] Nordiazepam (100ng/ml) [30-800ng/ml] Oxazepam (100ng/ml) [30-800ng/ml] Phenazepam (100ng/ml) [30-800ng/ml]	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS



Accredited to
ISO/IEC 17025:2017

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,
Trading as Randox Testing Services**

Issue No: 041 Issue date: 23 January 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/Equipment/Techniques used
BODY FLUIDS (cont'd) Urine (Preserved and Unreserved)	<p><u>Workplace Drug Analysis and Medical Legal Analysis (cont'd)</u></p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Benzodiazepines (cont'd):- Temazepam (100ng/ml) [30-800ng/ml]</p> <p>Gabapentinoids:- Pregabalin (100ng/ml) [30-800ng/ml] Gabapentin (100ng/ml) [30-800ng/ml]</p> <p>Ergoline:- LSD (1ng/ml) [0.3-8.0ng/ml]</p> <p>Phencyclidine:- PCP (25ng/ml) [3-80ng/ml]</p> <p>Quinazoline:- Methaqualone (100ng/ml) [30-800ng/ml]</p>	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS

END