

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> TESTING</p> <p>4348</p> <p>Accredited to ISO/IEC 17025:2017</p>	<p><b>Randox Laboratories Limited,</b> <b>Trading as Randox Testing Services</b></p> <p>Issue No: 042    Issue date: 11 March 2026</p>	
	<p>34 Diamond Road Crumlin Co Antrim BT29 4QY</p>	<p>Contact: Ms Louise Boyle Tel: +44 (0)2894 422413 Fax: +44 (0)1617 412777 E-Mail: Louise.Boyle@randox.com Website: www.randoxtestingservices.com</p>
<p><b>Testing performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS</p> <p>Urine (Preserved and unpreserved)</p>	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u></p> <p>Presumptive Screening for the presence of drugs of abuse (DOAI and DOA II) (cut-off limit):                      Amphetamines (500 ng/mL)                      Methamphetamine (500 ng/mL)                      Benzodiazepines (200 ng/mL)                      Barbiturates (200 ng/mL)                      Opiates (300 ng/mL)                      Cocaine metabolites (150 ng/mL)                      Methadone (300 ng/mL)                      Cannabinoids (50 ng/mL)                      Phencyclidine (25 ng/mL)                      Ketamine (75ng/ml)                      Oxycodone Class 1 and Class 2 (100ng/ml)                      Generic opioids (100ng/ml)                      Fentanyl (1ng/ml)                      Propoxyphene (300ng/ml)                      Methaqualone (300ng/ml)</p>	<p>Documented in house method SOP RTL 2213 RTL using Randox Evidence Investigator (method reference EI) biochip assay</p>



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,**  
**Trading as Randox Testing Services**

**Issue No: 042    Issue date: 11 March 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUIDS (cont'd)	<u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)	
Urine (Preserved and unpreserved)	Presumptive Screening for the presence of drugs of abuse (cont'd) (DOAI and DOA II) (cut-off limit): Buprenorphine (2ng/ml) LSD (0.5ng/ml) MDMA Group (500ng/ml)	Documented in house method SOP RTL 2213 RTL using Randox Evidence Investigator (method reference EI) biochip assay
Urine (Preserved and unpreserved)	Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:  $\Delta$ 9-Tetrahydrocannabinolic acid (THC-COOH) (15ng/ml) [5-150ng/ml]	Documented in house method A4URN and WI0227-RTL using solid phase extraction LC-MS/MS
Urine (Preserved and Unreserved)	Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:  Amphetamines:- Amphetamine (200ng/ml) [30-800ng/ml] Methamphetamine (200ng/ml) [30-800ng/ml] MDMA (200ng/ml) [30-800ng/ml] MDA (200ng/ml) [30-800ng/ml]	Documented in house method WI1099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS



4348

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Randox Laboratories Limited, Trading as Randox Testing Services

Issue No: 042 Issue date: 11 March 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUIDS (cont'd)  Urine (Preserved and Unreserved)	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Cocainics:- Benzoyllecgonine (100ng/ml) [30-800ng/ml] Cocaine (100ng/ml) [30-800ng/ml]</p> <p>Opiates:- 6-monoacetylmorphine (10ng/ml) [3-80ng/ml] Morphine (300ng/ml) [30-800ng/ml] Codeine (300ng/ml) [30-800ng/ml] Dihydrocodeine (300ng/ml) [30-800ng/ml]</p> <p>Opioids:- Methadone (250ng/ml) [30-800ng/ml] EDDP (75ng/ml) [30-800ng/ml] Oxycodone (100ng/ml) [30-800ng/ml] Oxymorphone (100ng/ml) [30-800ng/ml]</p>	<p>Documented in house method W11099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS</p>



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,**  
**Trading as Randox Testing Services**

**Issue No: 042    Issue date: 11 March 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS (cont'd)</p> <p>Urine (Preserved and Unreserved)</p>	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Opioids (cont'd):- Hydrocodone (100ng/ml) [30-800ng/ml] Hydromorphone (100ng/ml) [30-800ng/ml] Fentanyl (1ng/ml) [0.3-8.0ng/ml] Tramadol (100ng/ml) [30-800ng/ml]</p> <p>z-Drugs:- Zolpidem (10ng/ml) [3-80ng/ml]</p> <p>Ketamine:- Ketamine (75ng/ml) [30-800ng/ml] Norketamine (75ng/ml) [30-800ng/ml]</p>	<p>Documented in house method W11099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS</p>



4348

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Randox Laboratories Limited, Trading as Randox Testing Services

Issue No: 042 Issue date: 11 March 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
BODY FLUIDS (cont'd)  Urine (Preserved and Unreserved)	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Benzodiazepines:- Alpha-hydroxyalprazolam (100ng/ml) [30-800ng/ml] Diazepam (100ng/ml) [30-800ng/ml] Lorazepam (100ng/ml) [30-800ng/ml] Nordiazepam (100ng/ml) [30-800ng/ml] Oxazepam (100ng/ml) [30-800ng/ml] Temazepam (100ng/ml) [30-800ng/ml]</p> <p>Gabapentinoids:- Pregabalin (100ng/ml) [30-800ng/ml] Gabapentin (100ng/ml) [30-800ng/ml]</p> <p>Ergoline:- LSD (1ng/ml) [0.3-8.0ng/ml]</p> <p>Phencyclidine:- PCP (25ng/ml) [3-80ng/ml]</p>	<p>Documented in house method W11099-RTL using enzymatic hydrolysis (where applicable) and LC-MS/MS</p>



4348

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Randox Laboratories Limited, Trading as Randox Testing Services

Issue No: 042 Issue date: 11 March 2026

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS and TISSUES</p> <p>Human Hair (Head and Body)</p>	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Amphetamine (0.2ng/mg) [0.1-10ng/mg]</p> <p>Methamphetamine (0.2ng/mg) [0.1-10ng/mg]</p> <p>MDMA (0.2ng/mg) [0.1-10ng/mg]</p> <p>MDA (0.2ng/mg) [0.1-10ng/mg]</p> <p>Cocaine (0.5ng/mg) [0.25-25ng/mg]</p> <p>Morphine (0.1-10ng/mg) [0.2ng/mg]</p> <p>6-MAM (0.2ng/mg) [0.1-10ng/mg]</p> <p>Codeine (0.2ng/mg) [0.1-10ng/mg]</p> <p>Dihydrocodeine (0.2ng/mg) [0.1-10ng/mg]</p> <p>Alprazolam (0.05ng/mg) [0.025-2.5ng/mg]</p> <p>Tramadol (0.2ng/mg) [0.1-10ng/mg]</p>	<p>Documented in house method HDT and WI1790-RTL using LC-MS/MS</p>



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,**  
**Trading as Randox Testing Services**

**Issue No: 042    Issue date: 11 March 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS and TISSUES (cont'd)</p> <p>Human Hair (Head and Body)</p>	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Oxazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Bromazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Clonazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Diazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Nordiazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Lorazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Nitrazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Phenazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Temazepam (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Chlordiazepoxide (0.05ng/mg) [0.025-1.0ng/mg]</p> <p>Methadone (0.2ng/mg) [0.1-10ng/mg]</p> <p>EDDP (0.05ng/mg) [0.025-1.0ng/mg]</p>	<p>Documented in house method HDT and WI1790-RTL using LC-MS/MS</p>



Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Randox Laboratories Limited,**  
**Trading as Randox Testing Services**

**Issue No: 042    Issue date: 11 March 2026**

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<p>BODY FLUIDS and TISSUES (cont'd)</p> <p>Human Hair (Head and Body)</p>	<p><u>Workplace Drug Analysis and Medical Legal Analysis</u> (cont'd)</p> <p>Confirmation (above specified cut-off limit) and Quantitative analysis of the following drugs (cut-off limit) [concentration range]:</p> <p>Ketamine (0.5ng/mg) [0.25-25.0ng/mg] Norketamine (0.1ng/mg) [0.05-2.0ng/mg] THC (0.05ng/mg) [0.025-1.0ng/mg]</p>	<p>Documented in house method HDT and WI1790-RTL using LC-MS/MS</p>

END