Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK



Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details		Activity	Location code
Address Intech Calibration Ltd. Old Customs House Wharf Road Littlehampton West Sussex BN17 5DD	Local contact Mr J du Plessis Tel: +44 (0)1903 772 859 Fax: +44 (0) 903 754 437 Email: info@intechcalibration.co.uk Website: www.intechcalibration.co.uk	Temperature, Humidity and Electrical Calibration	Lab

Site activities performed away from the locations listed above:

Location details	Activity	Location code
Customer Premises, e.g. Hospitals, Laboratories and Manufacturing & Processing Plants	Temperature, Humidity and Electrical Calibration	Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK		
	Intech Calibration Limited		
4372 Accredited to ISO/IEC 17025:2017	Issue No: 026 Issue date: 09 May 2025		
Calibration performed by the Organisation at the locations specified			

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
TEMPERATURE			Calibration by comparison with reference instruments	
Temperature indicators with sensors	-196 °C -196 °C	0.040 °C 0.080 °C	Liquid Nitrogen	Lab Site
	0 °C	0.030 °C	Ice point	Lab & Site
	-100 °C to +155 °C	0.050 °C	Calibration in metal block baths	Lab & Site
	-100 °C to +0 °C 0 °C to 232 °C 232 °C to 420 °C 420 °C to 650 °C	0.15 °C 0.10 °C 0.13 °C 0.24 °C	Sensors with suitable dimensions	Lab & Site
Temperature calibration in air chamber, for instruments such as dataloggers	-40 °C to 0 °C 0 °C to 22 °C 22 °C to 60 °C 60 °C to 90 °C 90 °C to 100 °C 100 °C to 180 °C	0.45 °C 0.08 °C 0.09 °C 0.12 °C 0.59 °C 1.00 °C		Lab
	0 °C to 22 °C 22 °C to 60 °C	0.19 °C 0.22 °C		Lab & Site
In-situ temperature calibration in air	Ambient temperature	0.70 °C	Uncertainty will depend on stability of the ambient conditions	Site
Temperature controlled fridges, freezers, incubators, ovens and	-90 °C to +600 °C	1.0 °C	Multiple point measurements	Site
including associated recorders, indicators and controllers	-90 °C to +155 °C 155 °C to 600 °C	0.30 °C 0.70 °C	Single point measurement	Site

Calibration and Measurement Capability (CMC)

	Schedule of Accreditation issued by United Kingdom Accreditation Service		
	Intech Calibration Limited		
4372 Accredited to ISO/IEC 17025:2017	Issue No: 026 Issue date: 09 May 2025		
Calibration performed by the Organisation at the locations specified			

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
HUMIDITY				
Dew point	-32 °C to +90 °C	0.14 °C	By comparison with dew- point hygrometer and Platinum Resistance Thermometers	Lab
Relative humidity	Example conditions		Corresponding to above dew-point and temperature uncertainties	Lab
	Temperature range 0 °C to 22 °C 5 %rh to 50 %rh 50 %rh to 95 %rh	0.08 °C 0.20 %rh to 0.70 %rh 0.70 %rh to 1.20 %rh		
	Temperature range 22 °C to 60 °C 5 %rh to 50 %rh 50 %rh to 95 %rh	0.09 °C 0.20 %rh to 0.50 %rh 0.50 %rh to 0.90 %rh		
	Temperature range 60 °C to 90 °C 5 %rh to 50 %rh 50 %rh to 96 %rh	0.12 °C 0.20 %rh to 0.40 %rh 0.40 %rh to 0.80 %rh		
	Temperature range 0 °C to 22 °C 5 %rh to 50 %rh 50 %rh to 95 %rh	0.19 °C 0.20 %rh to 1.80 %rh 1.80 %rh to 2.40 %rh	By comparison with reference instruments	Lab & Site
	Temperature range 22 °C to 60 °C 5 %rh to 50 %rh 50 %rh to 95 %rh	0.22 °C 0.20 %rh to 1.90 %rh 1.90 %rh to 2.10 %rh		
Humidity controlled chambers, including associated recorders, indicators and controllers	Temperature range 0 °C to 60 °C 5 %rh to 95 %rh	0.23 °C 2.0 %rh	Single point measurement	Site
	Temperature range 60 °C to 90 °C 5 %rh to 95 %rh	0.32 °C 2.44 %rh		
	Temperature range 0 °C to 90 °C 5 %rh to 95 %rh	0.39 °C 2.95 %rh	Multipoint measurements	
In-situ humidity calibration in air	Ambient humidity	3.0 %rh	Uncertainty will depend on stability of the ambient conditions	Site

	Schedule of Accreditation issued by United Kingdom Accreditation Service 2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK			
UKAS CALIBRATION 4372 Accredited to ISO/IEC 17025:2017	Intech Calibration Limited Issue No: 026 Issue date: 09 May 2025			
	Calibration performed by the Organ	isation at the locations	specified	
Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty (k = 2)	Remarks	Location Code
ELECTRICAL All electrical measurements, inclu transfer to laboratory reference st	ding temperature simulation are carried andards unless otherwise determined ir	out using the method of the remarks column.	direct comparison or	
DC Voltage	0 V to 25 mV 25 mV to 75 mV 75 mV to 30 V	19 μV 25 μV 13 mV	Measure	Lab & site
	0V to 25 mV 25 mV to 75 mV 75 mV to 20 V	18 μV 29 μV 15 mV	Source	
DC Current	0 A to 24 mA	15 µA	Measure and Source	
Resistance	0 Ω to 400 Ω 400 Ω to 4 kΩ	42 mΩ 300 mΩ	Measure 2,3 & 4 wire	
	5 Ω to 400 Ω 400 Ω to 4 kΩ	120 mΩ 400 mΩ	Source 2 wire	
TEMPERATURE MEASUREMENT BY ELECTRICAL SIMULATION				Lab & site
Temperature indicators Thermocouple Type K N R T J S B E	-200 °C to 1200 °C -0 °C to 1200 °C 0 °C to 1600 °C 0 °C to 390 °C 0 °C to 1000 °C 0 °C to 1000 °C 800 °C to 1600 °C 0 °C to 990 °C	0.53 °C 0.53 °C 0.76 °C 0.35 °C 0.46 °C 0.80 °C 1.4 °C 0.72 °C	Including reference junction	
PRT Indicators PT100	-200 °C to 500 °C	0.25 °C		
i emperature calibrators Thermocouple Type				
K N R T J S B E	-200 °C to 1200 °C -0 °C to 1200 °C 0 °C to 1600 °C 0 °C to 390 °C 0 °C to 1000 °C 0 °C to 1000 °C 800 °C to 1600 °C 0 °C to 990 °C	0.67 °C 0.57 °C 1.3 °C 0.38 °C 0.53 °C 1.2 °C 1.0 °C 0.75 °C	Including reference junction compensation	
PRT Indicators PT 100	-200 °C to 500 °C	0.29 °C		
Internal reference junction calibration	Ambient	0.12 °C		
END				



Appendix - Calibration and Measurement Capabilities

Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of k = 2. An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means $1.5 \times 0.01 \times q$, where q is the quantity value.

The notation Q[a, b] stands for the root-sum-square of the terms between brackets: Q[a, b] = $[a^2 + b^2]^{1/2}$