

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>4393</b>  Accredited to <b>ISO/IEC 17025:2017</b>	<b>Eurofins MGS Laboratories Ltd</b>	
	<b>Issue No:</b> 015 <b>Issue date:</b> 10 December 2024	
	<b>Unit 2 Merlin Park</b> <b>Airport Service Road</b> <b>Portsmouth</b> <b>Hampshire</b> <b>PO3 5FU</b>	<b>Contact: Ms Kim Morwood</b> <b>Tel: +44 (0)2393 277990</b> <b>Fax: +44 (0)2393 277991</b> <b>E-Mail: emgs@bpt.eurofinseu.com</b> <b>Website: www.eurofins.co.uk/bpt</b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
DISINFECTANTS, ANTISEPTICS, BIOCIDES, HANDWASHES, HANDRUBS	<u>Microbiological Quantitative Disinfectant Suspension Testing:</u>	Documented in-house Work Instructions (methods) based on BS EN standard methods in full or with customer specified parameters
	Bacterial quantitative suspension test	WIN-1000.048 based on BS EN 1040:2005
	Bacterial quantitative suspension test for food, industrial, domestic and institutional applications	WIN-1000.050 based on BS EN 1276:2019
	Bacterial spore quantitative suspension test for food, industrial, domestic and institutional applications	WIN-1000.052 based on BS EN 13704:2018
	Bacterial quantitative suspension test for medical instrument applications	WIN-1000.053 based on BS EN 13727:2012 +A2:2015
	Fungal quantitative suspension test for medical instrument applications	WIN-1000.054 based on BS EN 13624:2021
	Fungal quantitative suspension test	WIN-1000.49 based on BS EN 1275:2005
	Fungal quantitative suspension test for food, industrial, domestic and institutional applications	WIN-1000.051 based on BS EN 1650:2019
	Bacterial quantitative suspension test for veterinary applications	WIN-1000.056 based on BS EN 1656:2019
	Fungal quantitative suspension test for veterinary applications	WIN-1000.057 based on BS EN 1657:2016



4393  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Eurofins MGS Laboratories Ltd**  
**Issue No:** 015    **Issue date:** 10 December 2024

Testing performed at main address only

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
DISINFECTANTS, ANTISEPTICS, BIOCIDES, HANDWASHES, HANDRUBS (cont'd)	<u>Microbiological Quantitative Handwash Disinfectant Testing:</u>  Bactericidal efficacy of hygienic handwashes  Bactericidal efficacy of hygienic handrubs  <u>Microbiological Quantitative Non-Porous Surface Disinfectant Testing:</u>  Quantitative surface test for food, industrial, domestic and institutional applications (non mechanical)  Bacterial and Fungal quantitative surface test for medical applications (mechanical employing impregnated wipes)  <u>Microbiological Quantitative Carrier Disinfectant Testing:</u>  Bacterial quantitative carrier test for medical instrument applications  Fungal quantitative carrier test for medical instrument applications	Documented in-house Work Instructions (methods) based on BS EN standard methods in full or with customer specified parameters  Following BS EN 1499:2013 supported by WIN-1000.063  Following BS EN 1500:2013 supported by WIN-1000.018  WIN-1000.058 based on BS EN 13697:2023  WIN-1000.130 based on EN 16615:2015  WIN-1000.060 based on EN 14561:2006  WIN-1000.061 based on BS EN 14562:2006
END		