


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>4481</b> <b>Accredited to ISO/IEC 17025: 2017</b>	<b>Water Research Centre Ltd</b>	
	<b>Issue No:</b> 014 <b>Issue date:</b> 17 October 2024	
	<b>Frankland Road Blagrove Swindon Wiltshire SN5 8YF</b>	<b>Contact: Liz Smith Tel: +44 (0)1793 865000 Fax: +44 (0)1793 865001 E-Mail: <a href="mailto:liz.smith@wrcgroup.com">liz.smith@wrcgroup.com</a> Website: <a href="https://www.wrcgroup.com/">https://www.wrcgroup.com/</a></b>
<b>Testing performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Materials/Products tested	Type of test/Properties measured/Range of measurement	Standard specifications/ Equipment/Techniques used
<b>SOLID WASTES</b>  Untreated and treated solid wastes from waste treatment and feedstocks to and outputs from anaerobic digestion  Untreated and treated solid wastes from waste treatment and feedstocks to and outputs from anaerobic digestion and Incinerator Bottom Ash  Untreated and treated solid wastes from municipal waste  Untreated and treated solid wastes from municipal waste and Incinerator Bottom Ash	<u>Chemical and Physical Tests</u>  BMc test - Biodegradability under anaerobic conditions  Dry Matter at 105°C and LOI at 550°C (organic matter content and ash content)  DR4 test - Biodegradability under aerobic conditions  Bulk Dry Matter at 40°C	  Documented in-house method PROCWASTE 011    Documented in-house method PROCWASTE 005 based on BS EN 13039:2011    Documented in-house method PROCWASTE 010 based on ASTM D5975-17    Documented in-house method PROCWASTE 002
<b>END</b>		